



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

User's Manual and Instructions

cfPure® V2 Cell-Free DNA Extraction Kit

Catalog Number: K5011610-V2, K5011625-V2, K5011625MA-V2

Storage Conditions

Store all of the contents of this kit at room temperature

Shelf Life

1 year from the date of receipt under proper storage conditions

Features

- Non-toxic chemicals
- High Cell-Free DNA recovery
- Short and Scalable Protocol
- Fully automatable using KingFisher Flex Purification System and Robotic Liquid Handlers
- Purified DNA is suitable for NGS, PCR, Bisulfite sequencing, etc

Description

BioChain's cfPure® Cell Free DNA extraction kit allows for fast and efficient Cell Free DNA (cfDNA) isolation from plasma/serum samples with minimal genomic DNA contamination. The magnetic bead-based extraction protocol is ideally suited for use with robotic liquid handlers and King Fisher Flex Purification System. This kit may also be applied on Hamilton's Presto system with the appropriate program. DNA extracted using this kit is suitable for PCR, qPCR, next generation sequencing (NGS), and other applications.

Contents

All necessary reagents for cfDNA isolation from human plasma samples are provided. There are 3 package sizes for this kit, which contains sufficient reagents for isolating cfDNA from up to 100 ml, 250 ml, or 250 ml's of sample in 5 - 10 ml increments.

Quality Control

Each component has been tested for purity and efficacy.

Important Notes

Blood Collection: The cfPure® Cell-Free DNA extraction kit has been optimized for use with samples collected in Streck Cell-Free DNA BCT, EDTA tubes and Acid Dextrose Acid (ACD) tubes.

Starting Material: Both fresh and frozen plasma can be used with the Cell Free DNA isolation protocol. Fresh plasma, however, tends to have higher yields.

Quantification: Plasma will yield 1-100 ng of Cell-Free DNA per ml of plasma. Therefore, quantification by absorbance measurement (eg. Nanodrop) may not be sensitive enough to accurately determine yield. Instead, we suggest using Qubit™ dsDNA High Sensitivity Assay.

Recommendations for PCR: Due to the highly fragmented nature of the nucleic acids obtained from plasma, care should be taken in the design of primers. Cell free DNA tends to have a small size (~170bp). Therefore, PCR primers should be designed to produce amplicons of 150 bp or less. Given the

AMSBIO | www.amsbio.com | info@amsbio.com



UK & Rest of the World
184 Park Drive, Milton Park
Abingdon OX14 4SE, UK
T: +44 (0)1235 828 200
F: +44 (0) 1235 820 482



North America
1035 Cambridge Street,
Cambridge, MA 02141
T: +1 (617) 945-5033 or
T: +1 (800) 987-0985
F: +1 (617) 945-8218



Germany
Bockenheimer Landstr. 17/19
60325 Frankfurt/Main
T: +49 (0) 69 779099
F: +49 (0) 69 13376880



Switzerland
Centro Nord-Sud 2E
CH-6934 Bioggio-Lugano
T: +41(0) 91 604 55 22
F: +41(0) 91 605 17 85

low concentration of cfDNA in plasma taken from healthy individuals, 40 amplification cycles may be needed in some cases.

Streck Cell-Free DNA BCT Tube(s): Plasma from blood collected with Streck Cell-Free DNA BCT Tube(s) must go through a Proteinase K treatment prior to Cell-Free DNA isolation to ensure optimal yields. Forgoing Proteinase K treatment may decrease yields by 50%

Equipment and Reagents to be Supplied by User

- Pipettes
- Vortex-Genie 2 or similar vortexing mixer*
- Magnet stand for molecular applications (e.g. DynaMag™-15 or DynaMag™-2)
- 1.5 ml non-stick eppendorf tube(s)
- Fresh 100% EtOH

* Contact BioChain® Technical Service for additional recommendations for high throughput or automated mixing.

Prior to Initial Use

The Lysis/Binding and Wash Buffer are shipped as a concentrate. If precipitate is present in either solution, incubate solution at 37°C for 30 minutes. 100% EtOH must be added to Lysis/Binding Buffer solutions before the first use. Once EtOH is added, these buffers are stable for one year. Be sure to close the bottle tightly for long term storage.

For 100 ml kit (K5011610-V2)

- Add 23 ml of fresh 100% ethanol to each bottle of Lysis/Binding Buffer and mix by inverting gently

For 250 ml kits (K5011625-V2)

- Add 19 ml of fresh 100% ethanol to each bottle of Lysis/Binding Buffer and mix by inverting gently

For 250 ml Max kits (K5011625MA-V2)

- Add 19 ml of fresh 100% ethanol to each bottle of Lysis/Binding Buffer and mix by inverting gently
- Prepare fresh 80% ethanol solution prior to each extraction

Before starting the protocol, determine the amount of plasma to be used for extraction and calculate the amount of buffer and beads needed. Any amount from 100 µl to 10 ml of plasma can be used. Scale buffer and bead volumes accordingly using the table below.

Sample Protocol

Plasma	Lysis/Binding Buffer	Bead Solution*	Elution** volume	Tube(s) size
x (x=ml of plasma)	1.25x	0.015x	0.20x	n/a
5 ml	6.25 ml	75 µl	100 µl	15 ml or 50ml**
7 ml	8.75 ml	105 µl	140 µl	50 ml
10 ml	12.50 ml	150 µl	200 µl	50 ml

* Using a 50 ml tube(s) for 5 ml or more of plasma is recommended over a 15 ml tube(s). While a 15 ml tube(s) will work it may lead to slightly lower yields

** This is recommended elution volume to get maximum yield, a low elution volume protocol for 10 ml plasma sample is provided below.

Proteinase K Treatment

If samples were collected using a **Streck Cell-Free DNA BCT tube(s)**, Proteinase K treatment is required to ensure optimal yields. If blood was not collected with **Streck Cell-Free DNA BCT tube(s)** proceed to the Lysis/Binding step.

Plasma	Proteinase K	20% SDS Solution
x (x=ml of plasma)	0.015 x	0.050 x
5 ml	75 µl	250 µl
7 ml	105 µl	350 µl
10 ml	150 µl	500 µl

1. Add the appropriate amount of plasma to an appropriately sized tube(s)
2. Add 15 µl of Proteinase K (20 mg/ml) for every 1 ml of plasma used
3. Add 50 µl of 20% SDS solution for every 1 ml of plasma used
4. Mix by inverting gently 5 times
5. Incubate at 60°C for 20 minutes
6. After incubation place tube(s) on ice for 5 minutes to cool tube(s) to room temperature
7. Once tube(s) are at room temperature proceed to step 2 of the Lysis/Binding section

Lysis/ Binding

1. Add the appropriate amount of plasma to appropriately sized tube(s)
2. Add 1.25 ml of **cfPure Lysis/Binding Buffer** for every 1 ml of plasma used
3. Add 15 µl of **cfPure Magnetic Bead Solution** for every 1 ml of plasma
Important: Mix beads well prior to adding. There should be no visible sedimentation at the bottom of the solution after mixing.
4. Vortex or shake tube(s) vigorously for 10 minutes at room temperature
 * To obtain high yields, ensure that plasma/buffer solution is mixing vigorously in tube(s). A vortexing mixer with a tube(s)-holder that allows for walk-away mixing will make this easier.
5. Place tube(s) into a magnet stand for 2 to 5 minutes, or until solution clears
6. While keeping the tube(s) on the magnet stand, remove supernatant. Be careful not to remove magnetic particles
7. Keep tube(s) on magnet stand for 1 minute, and remove residual supernatant

First Wash

8. Add 1000 µl of **cfPure Wash Buffer** to lysis/binding tube(s)
9. Resuspend beads by vortexing for 10 seconds or pipetting up and down 10 times
10. Transfer magnetic particle suspension into 1.5 ml micro tube(s) on magnet stand
11. Allow beads to attach to magnet stand for 10-30 seconds
12. Pipette supernatant from 1.5 ml tube(s) and use the supernatant to wash the lysis/binding tube(s)
13. Transfer the rest of the magnetic particles in lysis/binding tube(s) to the 1.5 ml tube(s)
14. Keep tube(s) on magnet stand for 10-30 seconds or until solution is clear
15. Remove as much buffer as possible using a 1000 µl pipette
16. Tap magnet stand on bench 5 times and remove remaining wash buffer with 200 µl pipette
17. Transfer tube(s) to non-magnetic rack and add 1000 µl of **cfPure Wash Buffer**
18. Resuspend beads by vortexing for 20 seconds or pipetting up and down 10 times
19. Centrifuge tube(s) briefly

*Centrifuge steps are needed when using vortexing to resuspend beads. Only a brief spin is recommended to remove solution from tube(s) lid
20. Place tube(s) on magnet stand for 10-30 seconds
21. Remove as much buffer as possible using a 1000 µl pipette
22. Tap magnet stand on bench 5 times and remove remaining wash buffer with 200 µl pipette

Second Wash

23. Transfer tube(s) to non-magnetic rack and add 1000 µl of **80% EtOH**
24. Resuspend beads by vortexing for 20 seconds or pipetting up and down 10 times
25. Centrifuge tube(s) briefly
26. Place on magnet stand for 10-30 seconds or until solution clears
27. Remove as much buffer as possible using a 1000 µl pipette
28. Tap magnet stand on bench 5 times and remove remaining EtOH with 200 µl pipette
29. Transfer tube(s) to non-magnetic rack and add 1000 µl of **80% EtOH**
30. Resuspend beads by vortexing for 20 seconds or pipetting up and down 10 times
31. Centrifuge tube(s) briefly
32. Place on magnet stand for 10-20 seconds
33. Remove as much EtOH as possible using a 1000 µl pipette and leave cap open
34. Tap magnet stand with tube(s) on bench 5 times

35. Remove remaining EtOH with 200 µl pipette
36. Leave tube(s) open on magnet stand for two minutes and then tap tube(s) on bench 5 times and remove any remaining EtOH with 20 µl pipette
37. Allow magnetic particles to dry for an additional 1-3 minutes
*Be careful to not over dry or beads may stick to tube(s)

Elution Step

38. Transfer microtube(s) to non-magnetic rack and add desired volume of **cfPure Elution Buffer** and resuspend beads
Important:: A minimum of 20 µl of cfPure Elution Buffer per ml of plasma is recommended to elute DNA to ensure optimal yields
39. Vortex or shake tube(s) vigorously for 5 minutes
40. Centrifuge tube(s) briefly
41. Place tube(s) on magnetic rack for 10 to 30 seconds
42. Transfer elute into a new 1.5 ml tube(s)

Alternative Low Elution Volume Protocol

The low elution volume protocol is recommended when 50 μ l or less of elution when extracting from 10 ml samples.

Plasma	Lysis/Binding Buffer	Bead Solution	Tube(s) size
10 ml	12.50 ml	75 μ l	50 ml

Proteinase K Treatment

If samples were collected using a **Streck Cell-Free DNA BCT tube(s)**, Proteinase K treatment is required to ensure optimal yields. If blood was not collected with **Streck Cell-Free DNA BCT tube(s)** proceed to the Lysis/Binding step.

Plasma	Proteinase K	20% SDS Solution
x (x=ml of plasma)	0.015 x	0.050 x
10 ml	150 μ l	500 μ l

1. Add the 10 ml's of plasma to an appropriately sized tube(s)
2. Add 150 μ l of Proteinase K (20 mg/ml)
3. Add 500 μ l of 20% SDS solution
4. Mix by inverting gently 5 times
5. Incubate at 60°C for 20 minutes
6. After incubation place tube(s) on ice for 5 minutes to cool tube(s) to room temperature
7. Once tube(s) are at room temperature proceed to step 2 of the Lysis/Binding section

Lysis/ Binding

1. Add the 10 ml of plasma to 50 ml tube(s)
2. Add 12.5 ml of **cfPure Lysis/Binding Buffer** to the tube(s)
3. Add 75 μ l of **cfPure Magnetic Bead Solution** to the tube(s)
Important: Mix beads well prior to adding. There should be no visible sedimentation at the bottom of the solution after mixing.
4. Vortex or shake tube(s) vigorously for 10 minutes at room temperature
 * To obtain high yields, ensure that plasma/buffer solution is mixing vigorously in tube(s). A vortexing mixer with a tube(s)-holder that allows for walk-away mixing will make this easier.
5. Place tube(s) into a magnet stand for 5 to 10 minutes, or until solution clears

6. While keeping the tube(s) on the magnet stand, remove supernatant. Be careful not to remove magnetic particles
7. Keep tube(s) on magnet stand for 1 minute, and remove residual supernatant

First Wash

8. Add 1000 µl of **cfPure Wash Buffer** to lysis/binding tube(s)
9. Resuspend beads by swirling tube or pipetting up and down 10 times
10. Transfer magnetic particle suspension into 2 ml micro tube(s) on magnet stand
11. Allow beads to attach to magnet stand for 20-30 seconds or until solution clears
12. Pipette supernatant from 2 ml tube(s) and use the supernatant to wash the lysis/binding tube(s)
13. Transfer the rest of the magnetic particles in lysis/binding tube(s) to the 2 ml tube(s)
14. Keep tube(s) on magnet stand for 20-30 seconds or until solution is clear
15. Remove as much buffer as possible using a 1000 µl pipette
16. Tap magnet stand on bench 5 times and remove remaining wash buffer with 200 µl pipette
17. Transfer tube(s) to non-magnetic rack and add 1000 µl of **cfPure Wash Buffer**
18. Resuspend beads by vortexing for 25 seconds or pipetting up and down 10 times
19. Centrifuge tube(s) briefly

*Centrifuge steps are needed when using vortexing to resuspend beads. Only a brief spin is recommended to remove solution from tube(s) lid
20. Place tube(s) on magnet stand for 20-30 seconds or until solution clears
21. Remove as much buffer as possible using a 1000 µl pipette
22. Tap magnet stand on bench 5 times and remove remaining wash buffer with 200 µl pipette

Second Wash

23. Transfer tube(s) to non-magnetic rack and add 1000 µl of **80% EtOH**
24. Resuspend beads by vortexing for 20 seconds or pipetting up and down 10 times
25. Centrifuge tube(s) briefly
26. Place on magnet stand for 20-30 seconds or until solution clears
27. Remove as much buffer as possible using a 1000 µl pipette
28. Tap magnet stand on bench 5 times and remove remaining EtOH with 200 µl pipette
29. Transfer tube(s) to non-magnetic rack and add 1000 µl of **80% EtOH**
30. Resuspend beads by vortexing for 20 seconds or pipetting up and down 10 times
31. Centrifuge tube(s) briefly
32. Place on magnet stand for 20-30 seconds or until solution clears

33. Remove as much EtOH as possible using a 1000 µl pipette and leave cap open
34. Tap magnet stand with tube(s) on bench 5 times
35. Remove remaining EtOH with 200 µl pipette
36. Leave tube(s) open on magnet stand for two minutes and then tap tube(s) on bench 5 times and remove any remaining EtOH with 20 µl pipette
37. Allow magnetic particles to dry for an additional 4 minutes

*Be careful to not over dry or beads may stick to tube(s)

Elution Step

38. Transfer microtube(s) to non-magnetic rack and add 50 ul of **cfPure Elution Buffer** and resuspend beads
39. Vortex or shake tube(s) vigorously for 5 minutes
40. Centrifuge tube(s) briefly
41. Place tube(s) on magnetic rack for 20 to 30 seconds or until solution clears
42. Transfer eluate into a new 1.5 ml tube(s)

Kit Components cfPure® V2 Cell-Free DNA , 100ml Kit (K5011610-V2)

Item	Cat#	Amount	Storage
1. Lysis/Binding Buffer	K5011610-V2-1	1 x 115 ml	Room Temp
2. Wash Buffer	K5011610-V2-2	2 x 105 ml	Room Temp
3. Elution Buffer	K5011610-V2-3	1 x 6 ml	Room Temp
4. Magnetic Bead Solution	K5011610-V2-4	2 x 1.33 ml	Room Temp

Kit Components cfPure® V2 cfDNA Extraction, 250ml Kit (K5011625-V2)

Item	Cat#	Amount	Storage
1. Lysis/Binding Buffer	K5011625-V2-1	3 x 95 ml	Room Temp
2. Wash Buffer	K5011625-V2-2	5 x 105 ml	Room Temp
3. Elution Buffer	K5011625-V2-3	1 x 15 ml	Room Temp
4. Magnetic Bead Solution	K5011625-V2-4	5 x 1.33 ml	Room Temp

Kit Components cfPure® MAX V2 cfDNA Extraction, 250ml Kit (K5011625MA-V2)

Item	Cat#	Amount	Storage
1. Lysis/Binding Buffer	K5011625MA-V2-1	3 x 95 ml	Room Temp
2. Wash Buffer	K5011625MA-V2-2	1 x 125 ml	Room Temp
3. Elution Buffer	K5011625MA-V2-3	1 x 15 ml	Room Temp
4. Magnetic Bead Solution	K5011625MA-V2-4	5 x 1.33 ml	Room Temp

cfPure® V2 Cell-Free DNA Extraction Kit

Isolation of cfDNA from 1 - 5 ml of sample using KingFisher™ Flex Magnetic Processor 24DW

Catalog Number: K5011610-V2, K5011625-V2

Product Description

Biochain's new cfPure Cell-Free DNA Extraction Kit has been designed to isolate circulating DNA from human plasma and serum. The cfPure utilizes Silica coated magnetic bead technology allowing for scalability and easy automation. The cfPure Cell-Free DNA extraction Kit can be used to isolate cfDNA from up to 24 samples of 1 - 5 ml of plasma or serum using the KingFisher™ Flex Magnetic Processor with 24 Deep Well Head. This guide describes the use of the cfPure kit with the KingFisher™ Flex Magnetic Processor 24DW to process samples of 1 - 5 ml.

Kit Contents and Storage

cfPure cfDNA Extraction Kit, 100ml (K5011610-V2)

Item	Amount	Storage
cfPure Lysis/Binding Buffer	1 x 115 ml	Room Temp.
cfPure Wash Buffer	2 x 105 ml	
cfPure Elution Buffer	1 x 6 ml	
cfPure Magnetic Bead Solution	2 x 1.33 ml	

King Fisher Flex 24 Deep Well Tip Comb and Plate	Thermofisher 97002610
--	-----------------------

cfPure cfDNA Extraction Kit, 250ml (K5011625-V2)

Item	Amount	Storage
cfPure Lysis/Binding Buffer	3 x 95 ml	Room Temp.
cfPure Wash Buffer	5 x 105 ml	
cfPure Elution Buffer	1 x 15 ml	
cfPure Magnetic Bead Solution	5 x 1.33 ml	

Item	Source
Consumables	
Aerosol-resistant pipette tips	Any
Nonstick, RNase-free Microfuge tubes (1.5ml)	Any
MicroAmp™ Clear Adhesive Film	Any
Reagent Reservoirs	Any
Reagents	
Ethanol, 200 proof (Absolute)	Any
SDS, 20% Solution (Only required for Proteinase K treatment)	Any
Proteinase K solution (20mg/ml) (Only required for Proteinase K treatment)	BioChain Z5050002

Equipment and Reagents to be Supplied by User

Item	Source
Equipment	
Multi-channel micropipettors	Any
Adjustable Micropipettors	Any
Vortexor	Any
Magnetic Particle Processor	
KingFisher™ Flex Magnetic Particle Processor	Thermofisher 5400630
Magnetic Head	
24 Deep-Well Plates for KingFisher™ Flex Magnetic Particle Processor	Thermofisher 24074440
Deep-Well Plates	
KingFisher™ Flex 24 deep well plate, sterile	Thermofisher 95040490
Tip Combs	

Download KingFisher™ Flex Program

1. On cfPure Webpage scroll down to Manual Section.
2. Click cfPure_4-5ml_Flex and/or cfPure_1-2ml_Flex to download program to your computer
3. Refer to KingFisher™ Flex manual for instructions for installing program on the instrument

Important Notes

Starting Material: Both fresh and frozen plasma can be used with the Cell-Free DNA isolation protocol. Fresh plasma, however, tends to have higher yields.

Streck Cell-Free DNA BCT Tube(s): Plasma from blood collected with Streck Cell-Free DNA BCT Tube(s) must go through a Proteinase K treatment prior to Cell-Free DNA isolation to ensure optimal yields. Forgoing Proteinase K treatment may decrease yields by 50%.

Protocol

Prior to Initial Use

The Lysis/Binding and Wash Buffer are shipped as a concentrate. If precipitate is present in either solution, incubate solution at 37°C for 30 minutes. 100% EtOH must be added to Lysis/Binding solution before the first use. Once EtOH is added, these buffers are stable for one year. Be sure to close the bottle tightly for long term storage.

For 100 ml kit (K5011610-V2)

- Add 23 ml of fresh 100% ethanol to each bottle of Lysis/Binding Buffer and mix by inverting gently
- Prepare fresh 80% ethanol solution prior to each extraction

For 250 ml kit (K5011625-V2)

- Add 19 ml of fresh 100% ethanol to each bottle of Lysis/Binding Buffer and mix by inverting gently
- Prepare fresh 80% ethanol solution prior to each extraction

Proteinase K Treatment

If samples were collected using a **Streck Cell-Free DNA BCT tube(s)**, Proteinase K treatment is required to ensure optimal yields. If blood was not collected with **Streck Cell-Free DNA BCT tube(s)** proceed to the Lysis/Binding step.

Plasma	Proteinase K	20% SDS Solution
x (x=μl of plasma)	0.015 x	0.050 x
1 ml	15 μl	50 μl
2 ml	30 μl	100 μl
4 ml	60 μl	200 μl
5 ml	75 μl	250 μl

8. Add the appropriate amount of plasma to an appropriately sized tube(s)
9. Add 15 μl of Proteinase K (20 mg/ml) for every 1 ml of plasma used
10. Add 50 μl of 20% SDS solution for every 1 ml of plasma used
11. Mix by inverting gently 5 times
12. Incubate at 60°C for 20 minutes
13. After incubation place tube(s) on ice for 5 minutes to cool tube(s) to room temperature

Plate Set up for 1 or 2 ml samples

Set up 24 Deep Well Plates by adding appropriate reagents according to table below

Plate ID	Plate Type	Plate Position	Reagent	Volume per well	
				1 ml	2 ml
Lysis/Binding Plate	24 DW Plate	1	cfPure Lysis/Binding Buffer	1.25 ml	2.5 ml
			cfPure Magnetic Bead Solution	15µl	30 µl
Wash Plate 1	24 DW Plate	2	cfPure Wash Buffer	1 ml	
Wash Plate 2	24 DW Plate	3	cfPure Wash Buffer	1 ml	
Wash Plate 3	24 DW Plate	4	80% Ethanol	2 ml	
Wash Plate 4	24 DW Plate	5	80% Ethanol	1 ml	
Elution Plate	24 DW Plate	6	cfPure Elution Buffer	50 - 100 µl	
Tip Comb	24 DW Plate	7	Place a 24 Deep-Well Tip Comb in Plate		

Plate Set up for 4 or 5 ml samples

Set up 24 Deep Well Plates by adding appropriate reagents according to table below

Plate ID	Plate Type	Plate Position	Reagent	Volume per well	
				4 ml	5 ml
Lysis/Binding Plate 1	24 DW Plate	1	cfPure Lysis/Binding Buffer	2.5 ml	3.125 ml
			cfPure Magnetic Bead Solution	30 µl	37.5 µl
Lysis/Binding Plate 2	24 DW Plate	2	cfPure Lysis/Binding Buffer	2.5 ml	3.125 ml
			cfPure Magnetic Bead Solution	30 µl	37.5 µl
Wash Plate 1	24 DW Plate	3	cfPure Wash Buffer	1 ml	
Wash Plate 2	24 DW Plate	4	cfPure Wash Buffer	1 ml	
Wash Plate 3	24 DW Plate	5	80% Ethanol	2 ml	
Wash Plate 4	24 DW Plate	6	80% Ethanol	1 ml	
Elution Plate	24 DW Plate	7	cfPure Elution Buffer	50 - 100 µl	
Tip Comb	24 DW Plate	8	Place a 24 Deep-Well Tip Comb in Plate		

- Gently shake Lysis/Binding Plate(s) to mix the reagents
- If extracting cfDNA from a 2 ml sample add entire sample to a well on Lysis/Binding Plate
- If extracting cfDNA from a 4 or 5 ml sample add half of sample to a well on Lysis/Binding Plate 1 and the other half of sample to the same well on Lysis/Binding Plate 2

Instrument Set up

- Place 24 Deep-Well magnetic head on to machine according the manuals protocol
- Select cfPure_4-5ml_Flex on the instrument for 4 or 5 ml extraction or cfPure_1-2ml_Flex for 1 or 2 ml extractions
- Start the run and follow on screen prompts to load processing plates in their respective positions
- At the end of the run remove elution plate from machine and cover plate or transfer eluate to new tubes

Isolated cfDNA is ready for immediate use or can be stored at -20°C

cfPure[®] V2 Cell-Free DNA Extraction Kit

Isolation of cfDNA using KingFisher[™] Flex Magnetic Processor 96DW

Catalog Number: K5011610-V2, K5011625-V2

Product Description

Biochain's new cfPure Cell-Free DNA Extraction Kit has been designed to isolate circulating DNA from human plasma and serum. The cfPure utilizes Silica coated magnetic bead technology allowing for scalability and easy automation. The cfPure Cell-Free DNA extraction Kit can be used to isolate cfDNA from up to 96 samples of 600 µl of plasma or serum using the KingFisher[™] Flex Magnetic Processor with 96 Deep Well Head. This guide describes the use of the cfPure kit with the KingFisher[™] Flex Magnetic Processor 96DW to process samples of 1000 µl or less.

Kit Contents and Storage

CfPure cfDNA Extraction Kit , 100ml (K5011610-V2)

Item	Amount	Storage
cfPure Lysis/Binding Buffer	1 x 115 ml	Room Temp.
cfPure Wash Buffer	2 x 105 ml	
cfPure Elution Buffer	1 x 6 ml	
cfPure Magnetic Bead Solution	2 x 1.33 ml	

96 Deep-Well Tip Combs for KingFisher [™] Flex Magnetic Particle	Thermofisher 97002534
---	-----------------------

CfPure cfDNA Extraction Kit , 250ml (K5011625-V2)

Item	Amount	Storage
cfPure Lysis/Binding Buffer	3 x 95 ml	Room Temp.
cfPure Wash Buffer	5 x 105 ml	
cfPure Elution Buffer	1 x 15 ml	
cfPure Magnetic Bead Solution	5 x 1.33 ml	

Item	Source
Consumables	
Aerosol-resistant pipette tips	Any
Nonstick, RNase-free Microfuge tubes (1.5ml)	Any
MicroAmp [™] Clear Adhesive Film	Any
Reagent Reservoirs	Any
Reagents	
Ethanol, 200 proof (Absolute)	Any
SDS, 20% Solution (Only required for Proteinase K treatment)	Any
Proteinase K solution (20mg/ml) (Only required for Proteinase K treatment)	BioChain Z5050002

Equipment and Reagents to be Supplied by User

Item	Source
Equipment	
Multi-channel micropipettors	Any
Adjustable Micropipettors	Any
Vortexor	Any
Magnetic Particle Processor	
KingFisher [™] Flex Magnetic Particle Processor 96DW	Thermofisher 5400630
Deep-Well Plates	
96 Deep-Well Plates for KingFisher [™] Flex Magnetic Particle Processor	Thermofisher 95040460
Standard Plates	
96 Standard Plates for KingFisher [™] Flex Magnetic Particle Processor	Thermofisher 97002540
Tip Combs	

Download KingFisher[™] Flex Program

1. On cfPure Webpage scroll down to Manual Section.
2. Click cfPure_600ul_Flex to download program to your computer
3. Refer to KingFisher[™] Flex manual for instructions for installing program on the instrument

Important Notes

Starting Material: Both fresh and frozen plasma can be used with the Cell-Free DNA isolation protocol. Fresh plasma, however, tends to have higher yields.

Streck Cell-Free DNA BCT Tube(s): Plasma from blood collected with Streck Cell-Free DNA BCT Tube(s) must go through a Proteinase K treatment prior to Cell-Free DNA isolation to ensure optimal yields. Forgoing Proteinase K treatment may decrease yields by 50%

Protocol

Prior to Initial Use

The Lysis/Binding and Wash Buffer are shipped as a concentrate. If precipitate is present in either solution, incubate solution at 37°C for 30 minutes. 100% EtOH must be added to Lysis/Binding solution before the first use. Once EtOH is added, these buffers are stable for one year. Be sure to close the bottle tightly for long term storage.

For 100 ml kit (K5011610-V2)

- Add 23 ml of fresh 100% ethanol to each bottle of Lysis/Binding Buffer and mix by inverting gently
- Prepare fresh 80% ethanol solution prior to each extraction

For 250 ml kit (K5011625-V2)

- Add 19 ml of fresh 100% ethanol to each bottle of Lysis/Binding Buffer and mix by inverting gently
- Prepare fresh 80% ethanol solution prior to each extraction

Proteinase K Treatment

If samples were collected using a **Streck Cell-Free DNA BCT tube(s)**, Proteinase K treatment is required to ensure optimal yields. If blood was not collected with **Streck Cell-Free DNA BCT tube(s)** proceed to the Lysis/Binding step.

Plasma	Proteinase K	20% SDS Solution
x (x=μl of plasma)	0.015 x	0.050 x
600 μl	9.0 μl	30 μl
1000 μl	15 μl	50 μl

14. Add the appropriate amount of plasma to an appropriately sized tube(s)
15. Add 15 μl of Proteinase K (20 mg/ml) for every 1 ml of plasma used
16. Add 50 μl of 20% SDS solution for every 1 ml of plasma used
17. Mix by inverting gently 5 times
18. Incubate at 60°C for 20 minutes
19. After incubation place tube(s) on ice for 5 minutes to cool tube(s) to room temperature

Plate Set up

Set up 96 Plates by adding appropriate reagents according to table below

	Plate Type	Plate Position on Instrument	Reagent	Volume per well	
				600 μ l	1000 μ l
Lysis/Binding Plate 1	96 Deep-Well Plate	1	cfPure Lysis/Binding Buffer	375 μ l	625 μ l
			cfPure Magnetic Bead Solution	4.5 μ l	7.5 μ l
Lysis/Binding Plate 2	96 Deep-Well Plate	2	cfPure Lysis/Binding Buffer	375 μ l	625 μ l
			cfPure Magnetic Bead Solution	4.5 μ l	7.5 μ l
Wash Plate 1	96 Deep-Well Plate	3	cfPure Wash Buffer	1 ml	1 ml
Wash Plate 2	96 Deep-Well Plate	4	cfPure Wash Buffer	1 ml	1 ml
Wash Plate 3	96 Deep-Well Plate	5	80% Ethanol	1 ml	1 ml
Wash Plate 4	96 Deep-Well Plate	6	80% Ethanol	500 μ l	500 μ l
Elution Plate	96 Standard Plate	7	cfPure Elution Buffer	30 - 50 μ l	30 - 50 μ l
Tip Comb	96 Deep-Well Plate	8		Place a 96 Deep-Well Tip Comb in Plate	

- Gently shake Lysis/Binding Plate 1 and 2 to mix the reagents
- Add half of plasma sample to the same wells of Lysis/Binding Plate 1 and 2

Instrument Set up

- Place 96 well Deep-Well magnetic head on to machine, and select cfPure_1ml_96_Flex on the instrument
- Start the run and follow on screen prompts to load processing plates in their respective positions
- At the end of the run remove elution plate and cover immediately

Isolated cfDNA is ready for immediate use

F-753-3UMRevD

K50116XX-V2 UB

Active Date: 02182020

AMSBIO | www.amsbio.com | info@amsbio.com



UK & Rest of the World
184 Park Drive, Milton Park
Abingdon OX14 4SE, UK
T: +44 (0)1235 828 200
F: +44 (0) 1235 820 482



North America
1035 Cambridge Street,
Cambridge, MA 02141
T: +1 (617) 945-5033 or
T: +1 (800) 987-0985
F: +1 (617) 945-8218



Germany
Bockenheimer Landstr. 17/19
60325 Frankfurt/Main
T: +49 (0) 69 779099
F: +49 (0) 69 13376880



Switzerland
Centro Nord-Sud 2E
CH-6934 Bioggio-Lugano
T: +41(0) 91 604 55 22
F: +41(0) 91 605 17 85