

Produktinformation



Forschungsprodukte & Biochemikalien
Zellkultur & Verbrauchsmaterial
Diagnostik & molekulare Diagnostik
Laborgeräte & Service

Weitere Information auf den folgenden Seiten! See the following pages for more information!



Lieferung & Zahlungsart siehe unsere Liefer- und Versandbedingungen

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien T. +43(0)1 489 3961-0 F. +43(0)1 489 3961-7 <u>mail@szabo-scandic.com</u> www.szabo-scandic.com

Pig IgG Heavy and Light Chain Antibody



Rabbit Polyclonal

Antigen Affinity Purified Catalog No. A100-114A

Lot No. 250106

LOU NO. 250	106
APPLICATIONS	WB, IHC, ICC, ELISA
SPECIES REACTIVITY	Pig
AMOUNT	1 ml
CONCENTRATION	1 mg/ml
STORAGE/SHELF LIFE	2 - 8°C / 1 year from date of receipt
PHYSICAL STATE	Liquid
BUFFER	Phosphate Buffered Saline (PBS) containing 0.09% Sodium Azide
ISOTYPE	IgG
ORIGIN	USA
PRODUCTION PROCEDURES	The antibody was isolated by affinity chromatography using antigen coupled to agarose beads.
	Immunoglobulin concentration was determined using Beer's Law where 1 mg/mL IgG has an A280 of 1.4.
	By immunoelectrophoresis and ELISA this antibody reacts specifically with pig IgG and with light chains common to other pig immunoglobulins. No antibody was detected against non- immunoglobulin serum proteins. This antibody may cross react with IgG from other species.
APPLICATIONS	Centrifuge tube to remove product from lid. Optimal working dilutions should be determined experimentally by the investigator. Prepare working dilution immediately before use.
	Western Blot1:1000 - 1:30,000Immunohistochemistry1:200 - 1:2000Immunocytochemistry1:200 - 1:2000ELISA1:1000 - 1:30,000; for coating plates 1:100 - 1:500
APPLICATION NOTES	Not all listed applications have been specifically tested by our laboratory.
ADDITIONAL INFO	https://www.fortislife.com/p/A100-114A Use the link above to view SDS, a current list of citations, and other product specific information.

This document certifies that this product has met all of the quality control standards defined by Bethyl Laboratories, Inc. Michael Spencer, PhD Date: January 7, 2025