



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 



6042 Cornerstone Court W, Ste B
San Diego, CA 92121
Tel: 1.858.202.1401
Fax: 1.858.481.8694
Email: info@bpsbioscience.com

Data Sheet

EZH1 Chemiluminescent Assay Kit

Catalog #52079

DESCRIPTION: The *EZH1 Chemiluminescent Assay Kit* is designed to measure activity of the EZH1 complex (EZH1/EED/SUZ12/RbAp48/AEBP) for screening and profiling purposes. The *EZH1 Assay Kit* comes in a convenient format, with a 96-well plate precoated with histone H3 peptide substrate, an antibody against methylated K27 residue of Histone H3, the secondary HRP-labeled antibody, S-adenosylmethionine, methyltransferase assay buffer, and purified EZH1 complex for 100 enzyme reactions. The key to the EZH1 Chemiluminescent Assay Kit is a highly specific antibody that recognizes methylated Histone H3K27. With this kit, only three simple steps on a microtiter plate are required for methyltransferase detection. First, S-adenosylmethionine is incubated with a sample containing assay buffer and methyltransferase enzyme for one hour. Next, primary antibody is added. Finally, the plate is treated with an HRP-labeled secondary antibody followed by addition of the HRP substrate to produce chemiluminescence that can then be measured using a chemiluminescence reader.

COMPONENTS:

Catalog #	Component	Amount	Storage	
51007	EZH1/EED/SUZ12/RbAp48/AEBP2	25 µg	-80°C	(Avoid freeze/thaw cycles!)
52120	400 µM S-adenosylmethionine	250 µl	-80°C	
52140F	Primary antibody 6	25 µl	-80°C	
52131H	Secondary HRP-labeled antibody 2	10 µl	-80°C	
52170	4x HMT assay buffer 2	3 ml	-20°C	
79556	Blocking buffer 1	50 ml	+4°C	
79670	ELISA ECL substrate A (translucent bottle)	6 ml	RT	
	ELISA ECL substrate B (brown bottle)	6 ml	RT	
	96-well plate precoated with histone substrate	1 plate	+4°C	

MATERIALS OR INSTRUMENTS REQUIRED BUT NOT SUPPLIED:

TBST buffer (1x TBS, pH 8.0, containing 0.05% Tween-20)

Luminometer or fluorescent microplate reader capable of reading chemiluminescence

Rotating or rocker platform

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.

To place your order, please contact us by Phone **1.858.202.1401** Fax **1.858.481.8694**

Or you can Email us at: info@bpsbioscience.com

Please visit our website at: www.bpsbioscience.com



6042 Cornerstone Court W, Ste B
San Diego, CA 92121
Tel: 1.858.202.1401
Fax: 1.858.481.8694
Email: info@bpsbioscience.com

APPLICATIONS: Great for studying enzyme kinetics and HTS applications.

STABILITY: One year from date of receipt when stored as directed.

REFERENCES:

Dillon SC, *et al. Genome Biol.* 2005; **6**:227.

Morin, RD., *et al. Nat. Genet.* 2010, **42**(2):181.

ASSAY PROTOCOL:

All samples and controls should be tested in duplicate.

Step 1:

- 1) Rehydrate the microwells by adding 150 μ l of TBST buffer (1x TBS, pH 8.0, containing 0.05% Tween-20) to every well. Incubate 15 minutes at room temperature. Tap the plate onto clean paper towels to remove liquid.
- 2) Thaw **S-adenosylmethionine** on ice. Upon first thaw, briefly spin tube containing **S-adenosylmethionine** to recover the full contents of the tube. Aliquot **S-adenosylmethionine** into single use aliquots. Store remaining **S-adenosylmethionine** in aliquots at -80°C immediately. *Note: S-adenosylmethionine is very sensitive to freeze/thaw cycles. Avoid multiple freeze-thaw cycles.*
- 3) Prepare the master mixture: N wells \times (7.5 μ l **4x HMT assay buffer 2** + 1.25 μ l **400 μ M S-adenosylmethionine** + 16.25 μ l H₂O)
- 4) Add 25 μ l of master mixture to each well designated for the "Positive Control", "Test Inhibitor", and "Blank". For the "Substrate Control", add 7.5 μ l **4x HMT assay buffer 2** + 17.5 μ l H₂O
- 5) Add 5 μ l of inhibitor solution of each well designated "Test Inhibitor". For the "Positive Control", "Substrate Control" and "Blank", add 5 μ l of the same solution without inhibitor (inhibitor buffer).
- 6) Prepare **1x HMT assay buffer** by diluting one part **4x HMT assay buffer 2** with three parts distilled water. Prepare only enough **1x HMT assay buffer** required for the assay. Add 20 μ l of **1x HMT assay buffer** to the well designated "Blank".

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.

To place your order, please contact us by Phone **1.858.202.1401** Fax **1.858.481.8694**

Or you can Email us at: info@bpsbioscience.com

Please visit our website at: www.bpsbioscience.com



6042 Cornerstone Court W, Ste B
San Diego, CA 92121
Tel: 1.858.202.1401
Fax: 1.858.481.8694
Email: info@bpsbioscience.com

	Blank	Substrate Control	Positive Control	Test Inhibitor
4x HMT assay buffer 2	7.5 µl	7.5 µl	7.5 µl	7.5 µl
400 µM S-adenosylmethionine	1.25 µl	–	1.25 µl	1.25 µl
H ₂ O	16.25 µl	17.5 µl	16.25 µl	16.25 µl
Test Inhibitor/Activator	–	–	–	5 µl
Inhibitor buffer (no inhibitor)	5 µl	5 µl	5 µl	–
1x HMT assay buffer	20 µl	–	–	–
EZH1 (12.5 ng/µl)	–	20 µl	20 µl	20 µl
Total	50 µl	50 µl	50 µl	50 µl

- 7) Thaw **EZH1** enzyme on ice. Upon first thaw, briefly spin tube containing enzyme to recover full content of the tube. Aliquot **EZH1** enzyme into single use aliquots. Store remaining undiluted enzyme in aliquots at -80°C immediately. *Note: **EZH1** enzyme is very sensitive to freeze/thaw cycles. Do not re-use thawed aliquots or diluted enzyme. Perform this step after all other dilutions and immediately prior to initiating reactions.*
- 8) Dilute **EZH1** enzyme in **1x HMT assay buffer** at 12.5 ng/µl (250 ng/20 µl). Keep diluted enzyme on ice until use. Discard any unused diluted enzyme after use.
- 9) Initiate reaction by adding 20 µl of diluted **EZH1** prepared as described above. Incubate at room temperature for 3 hours.
- 10) Wash the plate three times with 200 µl TBST buffer. Blot dry onto clean paper towels.
- 11) Add 100 µl of **Blocking buffer 1** to every well. Shake on a rotating platform for 10 min. Remove supernatant as above.

Step 2:

- 1) Dilute "**Primary antibody 6**" 400-fold with **Blocking buffer 1**.
- 2) Add 100 µl per well. Incubate 1 hour at room temperature with slow shaking.
- 3) Wash the plate three times with 200 µl TBST buffer. Blot dry onto clean paper towels.
- 4) Add 100 µl of **Blocking buffer 1** to every well. Shake on a rotating platform for 10 min. Remove supernatant as above.

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.

To place your order, please contact us by Phone **1.858.202.1401** Fax **1.858.481.8694**

Or you can Email us at: info@bpsbioscience.com

Please visit our website at: www.bpsbioscience.com



6042 Cornerstone Court W, Ste B
San Diego, CA 92121
Tel: 1.858.202.1401
Fax: 1.858.481.8694
Email: info@bpsbioscience.com

Step 3:

- 1) Dilute “**Secondary HRP-labeled antibody 2**” 1,000-fold with **Blocking buffer 1**.
- 2) Add 100 µl per well. Incubate for 30 min. at room temperature with slow shaking.
- 3) Wash the plate three times with 200 µl TBST buffer. Blot dry onto clean paper towels.
- 4) Add 100 µl of **Blocking buffer 1** to every well. Shake on a rotating platform for 10 min. Remove supernatant as above.
- 5) Just before use, mix on ice 50 µl **ELISA ECL substrate A** and 50 µl **ELISA ECL substrate B** and add 100 µl per well. Discard any unused chemiluminescent reagent after use.
- 6) Immediately read sample in a luminometer or microtiter-plate capable of reading chemiluminescence. “Blank” value is subtracted from all readings.

Reading Chemiluminescence:

Chemiluminescence is the emission of light (luminescence) which results from a chemical reaction. The detection of chemiluminescence requires no wavelength selection because the method used is emission photometry and is not emission spectrophotometry.

To properly read chemiluminescence, make sure the plate reader is set for LUMINESCENCE mode. Typical integration time is 1 second, delay after plate movement is 100 msec. Do not use a filter when measuring light emission. Typical settings for the Synergy 2 BioTek plate reader are: use the “hole” position on the filter wheel; Optics position: Top; Read type: endpoint. Sensitivity may be adjusted based on the luminescence of a control assay without enzyme (typically we set this value as 100).

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.

To place your order, please contact us by Phone **1.858.202.1401** Fax **1.858.481.8694**


Or you can Email us at: info@bpsbioscience.com

Please visit our website at: www.bpsbioscience.com



6042 Cornerstone Court W, Ste B
San Diego, CA 92121
Tel: 1.858.202.1401
Fax: 1.858.481.8694
Email: info@bpsbioscience.com

Example of Assay Results:



EZH1/EED/SUZ12/RbAp48/AEBP2 enzyme activity, measured using the EZH1 Chemiluminescent Assay Kit, BPS Bioscience Cat. #52079. Luminescence was measured using a Bio-Tek fluorescent microplate reader. *Data shown is lot-specific. For lot-specific information, please contact BPS Bioscience, Inc. at info@bpsbioscience.com.*

RELATED PRODUCTS

<u>Product</u>	<u>Cat. #</u>	<u>Size</u>
EZH1/EED/SUZ12/RbAp48/AEBP2	51007	50 µg
EZH1/EED/SUZ12/RbAp48/	51099	50 µg
EZH1/EED/SUZ12	51006	50 µg
EZH2/EED inactive	51002	20 µg
EZH2/EED/SUZ12	51003	50 µg
EZH2/EED/SUZ12/RbAp48/AEBP2	51004	50 µg
EZH2 (Y641F)/EED/SUZ12/RbAp48/AEBP2	51017	20 µg
EZH2 Homogeneous Assay Kit	52059	384 reactions
EZH2 Chemiluminescent Assay Kit	52009L	96 reactions
EZH2 (Y641F) Chemiluminescent Assay Kit	52075	96 reactions
EZH2 (A677G) Chemiluminescent Assay Kit	52077	96 reactions

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.

To place your order, please contact us by Phone **1.858.202.1401** Fax **1.858.481.8694**

Or you can Email us at: info@bpsbioscience.com

Please visit our website at: www.bpsbioscience.com



6042 Cornerstone Court W, Ste B
 San Diego, CA 92121
Tel: 1.858.202.1401
Fax: 1.858.481.8694
Email: info@bpsbioscience.com

TROUBLESHOOTING GUIDE

Problem	Possible Cause	Solution
Luminescence signal of positive control reaction is weak	EZH1 Complex has lost activity	Enzyme loses activity upon repeated freeze/thaw cycles. Use fresh EZH1 Complex, BPS Bioscience #51007. Store enzyme in single-use aliquots. Increase time of enzyme incubation. Increase enzyme concentration.
	Antibody reaction is insufficient	Increase time for primary antibody incubation. Avoid freeze/thaw cycles of antibodies.
	Incorrect settings on instruments	Refer to instrument instructions for settings to increase sensitivity of light detection.
	Chemiluminescent reagents mixed too soon	Chemiluminescent solution should be used within 15 minutes of mixing. Ensure both reagents are properly mixed.
Luminescent signal is erratic or varies widely among wells	Inaccurate pipetting/technique	Run duplicates of all reactions. Use a multichannel pipettor. Use master mixes to minimize errors.
	Bubbles in wells	Pipette slowly to avoid bubble formation. Tap plate lightly to disperse bubbles; be careful not to splash between wells.
Background (signal to noise ratio) is high	Insufficient washes	Increase number of washes. Increase wash volume. Increase Tween-20 concentration to 0.1% in TBST.
	Sample solvent is inhibiting the enzyme	Run negative control assay including solvent. Maintain DMSO level at <1% Increase time of enzyme incubation.
	Results are outside the linear range of the assay	Use different concentrations of EZH1 Complex, BPS Bioscience #51007 to create a standard curve.

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.

To place your order, please contact us by Phone **1.858.202.1401** Fax **1.858.481.8694**

Or you can Email us at: info@bpsbioscience.com

Please visit our website at: www.bpsbioscience.com