



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Corning Incorporated
Life Sciences

Registered
ISO 9001:2008

Product Description

Catalog Number: 1705-100

Product Description: Corning® Container, Sample, Coliform Water Test, Sterile, 100 mL, without Sodium Thiosulfate Tablet

Component Materials:
Polypropylene (PP)

Attributes:

Capacity - 120 mL total capacity. Capacity at line: 100 mL \pm 2.5 mL (2.5%)
Product Dimensions - 46.3 mm OD x 96 mm Height

Autoclaving:

This product is not designed to be autoclaved.

Heating / Cooling:

The intended use of this product does not involve heating or cooling.

The polypropylene material used to manufacture this product is rated for use from -196°C to 135°C.

Vacuum / Pressure:

This product is not designed for and is not rated for vacuum or pressure applications.

Sterilization:

This product has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products - Requirements for validation and routine control - Radiation sterilization*.

Sterility Assurance Level: SAL 10^{-3}

This product is bulk packed and therefore the exterior surfaces of the product should be assumed to be not sterile. The interior of the product is sterile and can be assumed to remain sterile until the cap is opened. The locking arrow feature between cap and container body is the visual indicator as to whether the cap has been opened. When the arrow feature is broken, it should be assumed that the cap has been opened and the original sterility of the interior has been compromised.

Lot Number Designation:

Each production lot is uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 2