



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Catalog Number: 3264

Product Description: Corning® CellCube® Culture Module, 100 Layer, TCT,

Fluid Path Component Materials:

Middle Plates	-	Virgin Polystyrene, meets USP Class VI requirements for plastic containers and closures.
End Plate	-	Polycarbonate, meets USP Class VI requirements for plastic containers and closures.
Encapsulate & Clips	-	Polycarbonate and white concentrate blend, both meet USP Class VI requirements for plastic containers and closures.
Hose Barb Adapter	-	Polycarbonate, meets USP Class VI requirements for plastic containers and closures.
Conical End Plug	-	Polycarbonate, meets USP Class VI requirements for plastic containers and closures.
Tubing	-	Silicone, meets USP Class VI requirements for plastic containers and closures.

Product Dimensions:

Overall Length	-	39 cm. (15.4 in)	Overall Width	-	25 cm. (9.8 in)
Overall Height	-	25 cm (9.8 in)	Tolerances	-	+/-1 cm. (+/-0.4 in)
Distance between plates	-	1 mm (reference dimension)			

Total CellCube Module:

85,000 cm²

Average Module Working Volume:

5.6 liters +/-0.2

Operating Pressures:

Not to exceed 5.0 psi

Sterilization – Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, “Sterilization of health care products- Radiation”. Products meet a minimum Sterility Assurance Level (SAL) of 10⁻⁶.

Sterility – Products labeled Sterile Fluid Path have been designed to ensure sterility of the portion of the product intended for contact with fluids.

Surface Characteristics – The surface is characterized to be hydrophilic and negatively charged, The negatively charged, carboxyl surface composition, has been optimized for cell attachment and growth.

BSE/TSE - Product complies with the latest revision of EMA/410/01 “Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products” by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01

Quality Control Testing – Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass

Packaging Inspection – Pass

Cell Attachment & Growth Treatment Verification – Pass

Leak Test – Pass

Catalog Number: 3264

Product Description: Corning® CellCube® Culture Module, 100 Layer, TCT,

Lot Number Designation:

8 Digit Lot Number: First 3 digits – Julian date, start of manufacturing; Next 2 digits – Year of manufacture; Last 3 digits – Batch identification.

Serial Number Designation:

12 Digit Serial Number; first 8 digits – Lot number (see Lot Number Designation above); Next 4 digits – Cellcube subassembly serial number.

Rev 12