



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Corning Incorporated Life Sciences

Registered
ISO 9001:2008

Product Description

Catalog Number: 3578

Product Description: Corning ® Spinner Flask, Disposable, 500ml, with vent caps

Component Materials:

- | | | |
|-------------|---|--|
| Vessel | - | Virgin Polystyrene, meets <i>USP, Class VI</i> requirements for plastic containers and closures, "complies with International Organization for Standardization (ISO) 10993-18, Biological Evaluation of Medical Devices - Part 18: Chemical Characterization of Materials (2005)." |
| Caps | - | High Density Polyethylene, meets <i>USP, Class VI</i> requirements for plastic containers and closures. |
| Filter | - | .2µm/polyester non-woven membrane, meets <i>USP, Class VI</i> requirements for plastic containers and closures. |
| Flange Base | - | Acrylonitrile Butadiene Styrene |
| Magnet | - | Polystyrene encapsulated ALNICO magnet, 12800 Gauss induction |

Product Dimensions:

Height of Vessel	-	8.0 in. (20.32cm)	Top Opening of Vessel	-	3.6 in. (9.21cm)
Diameter of Vessel	-	3.6in. (9.14cm)	Top Cap of Vessel	-	100mm GL 100
Footprint of Vessel	-	5.5 in. (13.97cm)	Sidearm Opening of Vessel	-	1.5 in. (3.81cm)
Sidearm Cap of Vessel	-	45mm GL 25	Height of Paddle	-	2.31 in. (5.87cm)
Width of Paddle	-	1.94 in. (4.92cm)	Working Volume	-	500ml minimum
Tolerances	-	+/- 0.05			

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (AAMI TIR 33) *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*. Sterility Assurance Level: SAL 10⁻⁶

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product is manufactured and packaged with animal free materials.

Pyrogens:

This vessel has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing*. Results: ≤ 0.1 EU/mL (≤ 4EU/device)

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 4