



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC Handels GmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Corning Incorporated Life Sciences

Registered
ISO 9001:2008

Product Description

Catalog Number: 3580

Product Description: Corning® Spinner Flask, Disposable, 1L, with Vent Caps

Component Materials:

- | | | |
|----------|---|--|
| Vessel | - | Virgin Polystyrene, <i>meets USP, Class VI</i> requirements for plastic containers and closures, "complies with International Organization for Standardization (ISO) 10993-18, Biological Evaluation of Medical Devices - Part 18: Chemical Characterization of Materials (2005)." |
| Impeller | - | Polypropylene, <i>meets USP, Class VI</i> requirements for plastic containers and closures, "complies with International Organization for Standardization (ISO) 10993-18, Biological Evaluation of Medical Devices - Part 18: Chemical Characterization of Materials (2005)." |
| Caps | - | High Density Polyethylene, <i>meets USP, Class VI</i> requirements for plastic containers and closures. Heavy metal free (meets CONEG requirements) color concentrate. |
| Filter | - | 0.2µm polyester non-woven membrane, <i>meets USP, Class VI</i> requirements for plastic containers and closures |
| Magnet | - | Polytetrafluoroethylene encapsulated ALNICO magnet, 12800 Gauss induction |

Product Dimensions:

- | | | | | | |
|---------------------------|---|--------------------|-----------------------|---|-------------------|
| Height of Vessel | - | 9.64 in. (24.48cm) | Diameter of Vessel | - | 8.0 in. (20.32cm) |
| Sidearm Opening of Vessel | - | 1.53 in (3.88cm) | Sidearm Cap of Vessel | - | 45 mm GL 45 |
| Working Volume | - | 1L maximum | Tolerances | - | +/- 0.05 |

Sterilization:

The product is irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (AAMI TIR 33) *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.
Sterility Assurance Level: SAL 10⁻⁶

Pyrogens:

This vessel has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing*.
Results: ≤ 0.1 EU/mL (≤ 4EU/device)

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product is manufactured and packaged with animal free materials.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 7