

Produktinformation



Forschungsprodukte & Biochemikalien
Zellkultur & Verbrauchsmaterial
Diagnostik & molekulare Diagnostik
Laborgeräte & Service

Weitere Information auf den folgenden Seiten! See the following pages for more information!



Lieferung & Zahlungsart siehe unsere Liefer- und Versandbedingungen

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien T. +43(0)1 489 3961-0 F. +43(0)1 489 3961-7 <u>mail@szabo-scandic.com</u> www.szabo-scandic.com

Corning Incor Life Scier ^{Registere} ISO 9001:2	Product Description
Catalog Number:	431219
Product Description:	Corning® syringe filter, surfactant free cellulose acetate (SFCA) membrane, 0.20 μ m pore size, 28 mm diameter, individually packaged
Component Materials: Housing-Membrane-	Acrylic Copolymer, meets <i>USP, Class VI</i> requirements for plastic containers and closures. Heavy metal free (meets <i>CONEG</i> req.) color concentrate. Surfactant Free Cellulose Acetate (SFCA), 0.20 μm pore size, meets <i>USP, Class VI</i> requirements for plastic containers and closures.
Product Dimensions: Diameter of unit Diameter of membrane Effective filtration area	 1.30 in. (33 mm) Length of syringe filter 1.10 in. (28 mm) Tolerances 4/- 0.05 in. (1.3 mm) 6.20 cm²

Sterilization:

This product has been irradiated and dosimetrically released based on the *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*. Sterility Assurance Level: SAL 10⁻⁵

Pyrogens:

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72: *Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing.* Results: less than 0.1EU/mL

Membrane Bubble Point:

The product has been tested according to an established procedure to determine the water bubble point of the product. The product meets the established criteria.

Housing Burst Pressure Test:

This product has been tested according to an established procedure to determine the housing burst pressure. The product meets the established criteria.

Flow Rate Test:

The product has been tested according to an established procedure to determine the flow rate. The product meets the established criteria.

Bacterial Challenge:

This product has been tested and determined to be quantitatively retentive when challenged with a concentration of 10⁷ organisms per cm² using *Brevundimonus diminuta*.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures. Pressure Hold Test: Forward pressurization of the unit to 1.5 bar. Visual Attributes: Visual examination of the product.

Rev No: 4