



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 431222

Product Description: Corning® syringe filter, regenerated cellulose (RC) membrane, 0.20 µM pore size, 25 mm diameter, individually packaged

Component Materials:

- Housing - Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.
- Membrane - Regenerated Cellulose (RC), 0.20 µM pore size, meets *USP, Class VI* requirements for plastic containers and closures.

Product Dimensions:

- | | | | | | |
|---------------------------|---|----------------------|----------------|---|-----------------------|
| Diameter of unit | - | 1.30 in. (33 mm) | Length of unit | - | 0.92 in. (23.3 mm) |
| Diameter of membrane | - | 0.98 in. (25 mm) | Tolerances | - | +/- 0.05 in. (1.3 mm) |
| Effective filtration area | - | 4.80 cm ² | | | |

Sterilization:

This lot has been irradiated and dosimetrically released based on the *Sterilization of healthcare products- Requirements for validation and routine control-Radiation sterilization.*

Sterility Assurance Level: SAL 10⁻⁵

Pyrogens:

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72: *Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing.* Results: less than 0.1EU/mL

HPLC Certified:

The product has been tested to ensure there are no extraneous peaks in the UV range.

Membrane Bubble Point:

The product has been tested according to an established procedure to determine the water bubble point of the product. The product meets the established criteria of >3.4 bar.

Housing Burst Pressure Test:

The product has been tested according to an established procedure to determine the housing burst pressure. The product meets the established criteria of >6 bar (approx. 87 psi).

Flow Rate Test:

The product has been tested according to an established procedure to determine the flow rate. The product meets the established criteria of >60 mL/min/bar.

Bacterial Challenge:

The product has been tested and determined to be quantitatively retentive when challenged with a concentration of 10⁷ organisms per cm² using *Brevundimonas diminuta*.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Pressure Hold Test: Forward pressurization of the unit to 1.0 bar.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date of manufacture; Next 2 digits - Year of manufacture; Last 3 digits - Run number for that year.

Rev No: 3