



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 431518

Product Description: 2L Erlenmeyer Flask w/Aseptic Connector: 2 Port, 1/4" Diptube, 0.2um Vent, 1/4" C-Flex w/Male MPC

Component Materials:

- Flask - Virgin Polycarbonate, meets USP, Class VI requirements for plastic containers and closures.
- Cap - Virgin Polypropylene, meets USP, Class VI requirements for plastic containers and closures. Heavy metal free (meets CONEG requirements) color concentrate.
- Tubing - C-flex thermoplastic elastomer, meets USP, Class VI requirements for Biological Test for Plastics.
- Filter - 50mm filter with 0.2 µm pore PVDF membrane, meets USP Biological Reactivity Test, In Vivo <88>
- Diptube - Polypropylene, meets USP, Class VI
- MPC Male - Medical Grade Polycarbonate, meets USP, Class VI
- Disconnect

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (TIR 33) Sterilization of Healthcare Products-Requirements for validation and routine control-Radiation sterilization.
Sterility Assurance Level: SAL 10-6

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

Performance Testing:

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

- | | |
|--------------------|---|
| Visual Attributes: | Visual examination of the product. |
| Packaging: | Inspection for seal and barrier integrity, accurate labeling and correct product configuration. |

Lot Number Designation:

8 Digit Lot Number: First 3 digits – Julian date, start of manufacturing; Next 2 digits – Year of manufacture; Last 3 digits – Batch identification.

Revision Date:

03-01-14

Rev No: 2