

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten! See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere Liefer- und Versandbedingungen

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

linkedin.com/company/szaboscandic in



Corning Incorporated
Life Sciences

Registered ISO 9001

Product description

Catalog Number 432331

Product Description: Corning® Octagonal PET Bottle, 125mL Graduated, 31mm IATA-Validated Screw Cap

Sterile, Pre-assembled, 24/Tray, 48/Case

Component Materials:

Bottle - PET, clear, non-cytotoxic, non-hemolytic, meets USP, Class VI requirements for plastic containers and

 Cap - HDPE, natural, non-cytotoxic, non-hemolytic, meets USP, Class VI requirements for plastic containers and closures.

Attributes:

Capacity - 125 mL; Tolerance: ±13 mL at any mark

Product Dimensions - 108 mm Height (with cap locked); 106 mm Height (w/o cap); 53 mm Width;

42 mm Cap OD; 31.7 mm Neck ID

Autoclaving:

This product is not suitable for autoclaving.

Heating / Cooling:

This product is designed for use at temperatures ranging from -40°C to 50°C.

Special Features:

The product is supplied with the cap in place but not locked.

This product is manufactured in an environment maintained to the standards of a Class 100 clean room.

Sterilization:

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10⁻⁶.

Leakproof:

With a torque value of 2,5N.m (22lb.in), this product meets IATA Packaging Instruction 650 requirements to ensure leakproof performance during air transport conditions; however, this is not designed for packaging and transport of UN3373 Biological and Infectious Substances.

Pyrogens:

This product is evaluated for pyrogens according to USP 31 and meets the requirements for classification as non-pyrogenic. Results less than 0.1 EU/mL.

Expiration Statement:

This product is defined to have and expiration date of 3 years from the date of manufacture.

Rev No: 1