



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Catalog Number: 4440

ProductDescription: Corning® Elplasia™ Microcavity Plate, 6-Well, Black Clear Round Bottom, 500/400, Ultra-Low Attachment with Lid

Component Materials:

Plate walls - Virgin Polystyrene, black concentrate
Plate bottom - Virgin Polystyrene
Lid - Virgin Polystyrene
Surface - Proprietary hydrogel

Product Dimensions:

Length of plate	-	5.03 in.	Working Volume	-	5mL
Width of plate	-	3.37 in.	Tolerances	-	+/- 0.05 in.
Height of plate with lid	-	0.91 in.			

Microcavity Nominal Dimensions:

Bottom diameter - 500µm
Height - 400µm

Sterilization:

The product has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*. Sterility Assurance Level: SAL 10-3

Pyrogens:

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72: *Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing*. Results: less than 0.1EU/mL.

Tissue Culture:

The product has been tested for the ability to promote the growth and formulation of 3D cellular spheroids of mammalian suspension cells in culture.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Revision Date:

03-17-20

Rev No: 2