



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated  
Life Sciences**

Registered  
ISO 9001:2008

**Product Description**

**Catalog Number:** 7000

**Product Description:** Costar® 100 mL Stripette®, individual, paper / plastic wrap

**Component Materials:**

- Pipette - Virgin Polystyrene, meets *USP Class VI requirements for plastic containers and closures*. Polystyrene also meets *USP Cytotoxicity Test <87>*.
- Mouthpiece Plug - Polyester, meets *USP Cytotoxicity Test <87>*.
- Ink - Aqua / black, meets *USP Cytotoxicity Test <87>*.

**Product Dimensions:**

- |                       |   |                           |               |   |         |
|-----------------------|---|---------------------------|---------------|---|---------|
| Overall Length        | - | 13.61 in. +/- .050 in     | Tip ID        | - | .125 in |
| Max Working Volume    | - | 100 mL                    | Mouthpiece OD | - | .310 in |
| Volumetric accuracy   | - | +/- 2% stated full volume | Pipette OD    | - | .960 in |
| Tolerances of ID & OD | - | +/- .015 in.              |               |   |         |

**Sterilization:**

The lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of health care products-Requirements for validation and routine control-Radiation sterilization*.  
Sterility Assurance Level: SAL 10<sup>-6</sup>

**RNase/DNase Testing:**

This product has been tested and is free of any detectable RNase/DNase contamination.

**Pyrogens:**

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 *Bacterial Endotoxins - Test methodologies, routine monitoring and alternative to batch testing*.  
Results: ≤ 0.1 EU/mL (≤ 4EU/device)

**Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:**

This product is manufactured and packaged with animal free materials.

**Performance Testing:**

- Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.
- Integrity Test: Forward pressurization of the product to a minimum of 1.5 psi.
- Visual Attributes: Visual examination of the product.
- Volumetric Accuracy: Gravimetric volume testing at stated full volume
- Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.