



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated  
Life Sciences**

Registered  
ISO 9001:2008

**Product Description**

**Catalog Number:** 7424

**Product Description:** Costar® Transwell®, 100 cm<sup>2</sup>, tissue culture treated, 0.4 µm pore size, MCP 1 0, with tray and lid

**Component Materials:**

- Tray / Lid - Virgin Polyethylene Terephthalate, 0.060 inches thick; meets *USP, Class VI* requirements for plastic containers and closures.
- Transwell Body - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
- Filter - Translucent polycarbonate PVP-free membrane, 0.4µm pore size, 10µm thick, meets *USP, Class VI* requirements for plastic containers and closures.
- Primary Packaging - Nylon, meets *USP, Class VI* requirements for plastic containers and closures.

**Product Dimensions:**

- |                          |   |           |  |   |           |
|--------------------------|---|-----------|--|---|-----------|
| Length of tray           | - | 6.927 in. | Bottom of transwell to tray (in high position) | - | 0.549 in. |
| Width of tray            | - | 4.803 in. | Bottom of transwell to tray (in low position)  | - | 0.124 in. |
| Height with Lid          | - | 1.462 in. | Volume added / tray (in high position)         | - | 197.7 mL  |
| Tolerances of Dimensions | - | +/- .050  | Volume added / tray (in low position)          | - | 42.2 mL   |

**Sterilization:**

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (AAMI TIR 33) *Sterilization of healthcare products - Requirements for validation and routine control-Radiation sterilization*.  
Sterility Assurance Level: SAL 10<sup>-6</sup>

**Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:**

This product is manufactured and packaged with animal free materials.

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Cell Culture Treatment: Wettability test using water to insure the presence of a hydrophilic surface.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 1