

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten! See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere Liefer- und Versandbedingungen

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

linkedin.com/company/szaboscandic in



Corning Incorporated Life Sciences

Registered ISO 9001:2008

Product Description

Catalog Number: 9903

Product Description: Corning® CellSTACK ®- 10 Chamber, with one piece universal cap and solid overcap

and solid overcap

Component Materials:

Top Plate - Virgin Polystyrene, meets USP, Class VI requirements for plastic containers and

closures.

Middle Plate - Virgin Polystyrene, meets USP, Class VI requirements for plastic containers and

closures.

Bottom Plate - Virgin Polystyrene, meets USP, Class VI requirements for plastic containers and

closures.

Adhesive - Proprietary Acrylate, meets ISO-10993, Biocompatibility requirements and does not

contain any animal products.

One Piece Universal

Cap

Virgin High Density Polyethylene, meets *USP*, *Class VI* requirements for plastic

containers and closures. Heavy metal free (meets CONEG requirements) color

concentrate.

Overcap - Dow LDPE 993I Natural

Product Dimensions:

Overall Length - 13.2 in. (335mm) Overall Width - 8.1 in. (206mm)

Overall Height of One - 8.771 in. (222.78mm)

Piece Universal Cap with

Overcap

Neck ID - 1.0 in. (26mm) Neck O.D. incl. Threads - 1.3 in. (32mm)
Tolerances - +/- 0.1 in. (2.5mm) Distance between plates - 0.67 in. (17mm)

Total Cell Growth Area:

6360 cm²

Recommended Working Volume:

1300 - 2000 ml

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (AAMI TIR33) *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*Sterility Assurance Level: SAL 10⁻⁶

Pyrogens:

The ČellStack has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing.

Results: < 0.1 EU/mL (< 4EU/device)

Surface Characterization:

Surface is characterized to be hydrophilic and negatively charged. The negatively charged, carboxyl surface composition, has been optimized for cell attachment and growth.

Cell Attachment and Growth Characteristics:

The product has been tested for the attribute of cell attachment and growth utilizing an attachment- dependent mammalian cell line in a serum supplemented media.

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product is manufactured and packaged with animal free materials

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Integrity Testing: Forward pressurization of the product to 1.5 psi.

Cell Culture Treatment: Wettability test using water to insure the presence of a hydrophilic

surface.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct

product configuration.

Corning Incorporated Life Sciences

Registered ISO 9001:2008

Product Description

Catalog Number: 9903

Product Description: Corning® CellSTACK ®- 10 Chamber, with one piece universal cap and solid overcap

and solid overcap

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Serial Number Designation:

12 Digit Serial Number: First 8 digits - Lot number (see lot number designation above); Next 3 digits CellSTACK serial number; Last digit - Stack designation.

Rev No: 11