



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 3801

Product Description: Costar® Snapwell™ 12mm, TC treated, 0.4 µm pore size, in a 6-well cluster plate with lid

Component Materials:

- Plate/Lid - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
- Transwell body - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
- Filter - Transparent polyester membrane, meets *USP, Class VI* requirements for plastic containers and closures.

Product Dimensions:

- | | | | | | |
|--------------------------|---|--------------|------------------------------|---|----------|
| Length of Plate | - | 5.030 in. | Height with Lid | - | .891 in. |
| Width of Plate | - | 3.365 in. | Bottom of Transwell to plate | - | .045 in. |
| Tolerances of Dimensions | - | +/- .050 in. | | | |

Special Feature:

The bottom section of the Snapwell is easily detachable from its upper assembly for placement in the Diffusion Chamber.

Sterilization:

The product is irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.

Sterility Assurance Level: SAL 10⁻³

Cell Attachment and Growth Characteristics:

The product has been tested for the attribute of cell attachment and growth utilizing an attachment-dependent mammalian cell line in a serum supplemented media.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Cell Culture Treatment: Wettability test using water to insure the presence of a hydrophilic surface.

Visual Attributes: Visual and microscopic examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 6