



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 4011

Product Description: Costar® 1 mL Stripette®, individual, plastic wrap

Component Materials:

- Pipette - Virgin Polystyrene, meets *USP Class VI requirements for plastic containers and closures*. Polystyrene also meets *USP Cytotoxicity Test <87>*.
- Mouthpiece Plug - Polyester, meets *USP Cytotoxicity Test <87>*.
- Ink - Yellow / black, meets *USP Cytotoxicity Test <87>*.

Product Dimensions:

- | | | | | | |
|-----------------------|---|---------------------------|---------------|---|----------|
| Overall Length | - | 10.94 in. +/- .050 in | Tip ID | - | .065 in |
| Max Working Volume | - | 1.2 mL | Mouthpiece OD | - | .180 in |
| Volumetric accuracy | - | +/- 2% stated full volume | Pipette OD | - | .180 in. |
| Tolerances of ID & OD | - | +/- .015 in. | | | |

Sterilization:

The lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of health care products-Requirements for validation and routine control-Radiation sterilization*.

Sterility Assurance Level: SAL 10⁻³

Pyrogens:

The product has been tested and has met the criteria established in ANSI/AAMI ST 72:2002/(R) 2010: *Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing*.

RNase/DNase Testing:

This product has been tested and is free of any detectable RNase/DNase contamination.

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product is manufactured and packaged with animal free materials.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Volumetric Accuracy:

Gravimetric volume testing at stated full volume.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 14