



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated  
Life Sciences**

Registered  
ISO 9001:2008

**Product Description**

**Catalog Number:** 4012

**Product Description:** Costar ® 1 mL Stripette ®, individual, plastic wrap

**Component Materials:**

- Pipette - Virgin Polystyrene, meets *USP Class VI requirements for plastic containers and closures*. Polystyrene also meets *USP Cytotoxicity Test <87>*.
- Mouthpiece Plug - Polyester, meets *USP Cytotoxicity Test <87>*.
- Ink - Yellow / black, meets *USP Cytotoxicity Test <87>*.

**Product Dimensions:**

- |                       |  |               |           |
|-----------------------|--|---------------|-----------|
| Overall Length        | - 10.94 in. +/- .050 in                | Tip ID        | - .065 in |
| Max Working Volume    | - 1.2 mL                               | Mouthpiece OD | - .180 in |
| Volumetric accuracy   | - +/- 2% stated full volume            | Pipette OD    | - .180 in |
| Tolerances of ID & OD | - Tolerances of ID & OD<br>+/- .015 in |               |           |

**Sterilization:**

The lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of health care products-Requirements for validation and routine control-Radiation sterilization*.

Sterility Assurance Level: SAL 10<sup>-3</sup>

**Pyrogens:**

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing*. Results: ≤ 0.1 EU/mL ( ≤ 4EU/device)

**RNase/DNase Testing:**

This product has been tested and is free of any detectable RNase/DNase contamination.

**Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:**

This product does not contain materials of animal origin.

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

- |                      |  |
|----------------------|--|
| Visual Attributes:   | Visual examination of the product.   |
| Volumetric Accuracy: | Gravimetric volume testing at stated full volume.  |
| Packaging:           | Inspection for seal and barrier integrity, accurate labeling, and correct product configuration. |

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 18