



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

## Product Description

**Catalog Number:** 6511

**Product Description:** Costar ® 96-well Thermowell™ plate, Model M

**Component Materials:**

Plate - Polycarbonate

**Product Dimensions:**

Length of Plate	-	5.039 in.	Diameter of Well @ top	-	.245 in.
Width of Plate	-	3.386 in.	Depth of Well	-	.500 in.
Tolerances of Dimensions	-	+/- .050 in.	Total Volume per well	-	200 µL

**Functional Characteristics:**

The product is thermally stable up to 135° C.

**Compatible Thermal Cyclers:**

MJ Research PTC-100-96V, PTC-200 DNA Engine™, Biometra Uno, Coy Corporation Temp Cycler II, Corbett Research FTS-960, Hybaid OmniGene with Microblock, Lab-Line Programmable Thermal Block, Quatro BioSystems, Stratagene Robocycler Gradient 96.

**Compatible Thermal Cycler:**

MJ Research PTC-100-96V, PTC-200 DNA Engine™, Biometra Uno, Coy Corporation Temp Cycler II, Corbett Research FTS-960, Hybaid OmniGene with Microblock, Lab-Line Programmable Thermal Block, Quatro BioSystems, Stratagene Robocycler Gradient 96.

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 3