



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

## ExoStep™

### Exosome FACS analysis for exosomes

REFERENCE	SIZE	DESCRIPTION
ExoS-25-G9	25 test	Includes CD9 biotin as primary detection antibody.
ExoS-50-G9	50 test	Includes CD9 biotin as primary detection antibody.
ExoS-25-G81	25 test	Includes CD81 biotin as primary detection antibody.
ExoS-50-G81	50 test	Includes CD81 biotin as primary detection antibody.

### INTRODUCTION

Exosomes are small extracellular vesicles that are released from cells upon fusion of an intermediate endocytic compartment, the multivesicular body (MVB)<sup>1</sup>, with the plasma membrane. They are thought to provide a means of intercellular communication<sup>(2,3)</sup> and of transmission of macromolecules between cells allowing the spread of proteins, lipids, mRNA, miRNA and DNA and as contributing factors in the development of several diseases. Exosomes can also modulate cancer microenvironment<sup>(4)</sup> and the immune response<sup>(5,6)</sup>.

### PRODUCT DESCRIPTION

The kit is a simple immunobead assay for isolation/detection of exosome, using a bead-bound anti-CD63 capture antibody and a fluorochrome conjugated anti-CD9/CD81 detection antibody. The kit provides reproducible results and can be run in parallel to exosome immunophenotyping.


Tested application: Flow Cytometry<sup>(7,8)</sup>.

Species reactivity: Human

Storage buffer: aqueous buffered solution containing protein stabilizer and 0.09% sodium azide (NaN<sub>3</sub>).

Recommended usage: Immunostep's ExoStep, is intended for the immunoisolation (immunomagnetic or FACS) and Flow Cytometry analysis of pre-enriched CD63<sup>+</sup> human exosomes from biofluids (plasma, urine) or cell culture media.

Presentation: liquid



**Figure 1:** Immunobeads assay for isolation/detection of exosome workflow.

### REAGENTS PROVIDED

DESCRIPTION	COMPONENTS	AMOUNT	Nº TEST
Superparamagnetic Capture Beads	CD63+ (Clone TEA3/18) beads. Polystyrene micro-particles with Mean Diameter (µm) 6.5±0.2 (CV<5%), having discrete fluorescence intensity characteristics	6000 beads/test (50 µl/test)	25/50 test
Primary detection antibody	Anti-CD9 biotin or Anti-CD81 biotin(Clone M38)	(5 µl/test)	25/50 test
Secondary detection reagent	Streptavidin-Phycoerythrin (PE) is useful for detecting biotinylated antibodies. The excitation of PE by 488nm laser light induces a light emission maximum of 575nm.	5 µl/test	25/50 test
Assay Buffer 10X	PBS 10% BSA, pH 7,4 – 10X . Do not freeze. Dilute contents of the 10X Assay Buffer to 1X (PBS 1% BSA) in PBS, for use in this assay.	10/15 ml	

### APPROPRIATE STORAGE AND HANDLING CONDITIONS

Store in the dark, refrigerated between 2 °C and 8 °C. DO NOT FREEZE. The kit is stable until the expiry date stated on the vial label if kept at 2-8°C. Do not use after the date indicated.

### REAGENTS NOT PROVIDED

- Pre-enriched exosomes by ultra-centrifugation.
- Magnetic Rack; MagneSphere(R) Mag. Sep. Stand 12-hole, 12x75mm (PROMEGA, Ref Z5343).
- 12x75mm Polystyrene Round Bottom Tubes (cytometer tubes).
- Sterile syringe filter with a 0.45µm pore (EMD Millipore Millex, Ref: SLHV033RS).
- Syringe of adequate volume.

### EVIDENCE OF DETERIORATION

Reagents should not be used if any evidence of deterioration is observed. For more information, please contact our technical service: tech@immunostep.com

### RECOMMENDATIONS AND WARNINGS



- Avoid microbial contamination of the reagent. Assay buffer 1X can be filtered before use.*
- Microspheres and reagents should be protected from prolonged exposure to light throughout this procedure.*
- The samples should be treated with appropriate handling procedures.*
- Depending on the type of exosomes used, the number of exosomes may vary with respect to the concentration of the protein.*
- Do not use after the expiry date indicated on the vial.*
- Deviations from the recommended procedure could invalidate the analysis results.*

- g) Before acquiring the samples, it is necessary to make sure that the flow cytometer is calibrated and compensated.
- h) The isolation and detection success is dependent on the quality of the sample pre-enrichment process.

## WARRANTY

Warranted only to conform to the quantity and contents stated on the label or in the product labelling at the time of delivery to the customer. Immunostep disclaims hereby other warranties.

Immunostep's sole liability is limited to either the replacement of the products or refund of the purchase price.


## SAMPLE PREPARATION

Exostep allows the detection of isolated exosomes from differential ultracentrifugation (a) as well as direct detection in the sample without the need for ultracentrifugation, just with simple pretreatment (b).

### a. Purification of Exosomes by Differential Ultracentrifugation.

The kit has been validated for pre-enriched human exosomes from cell culture and bodily fluids, such as serum/plasma, and urine, through an ultracentrifugation protocol (Fig. 2)<sup>9</sup>.

The principle for exosome purification is the same for cell culture and bodily fluids, but due to the viscosity of some fluids it is necessary to dilute them with an equal volume of PBS, before centrifugations.



**Figure 2:** Workflow for the exosome pre-enrichment based on differential ultracentrifugation.


Please, refer to <http://www.immunostep.com/content/34-protocols> for additional technical information.

### b. Sample pretreatment for direct exosome detection on cell culture supernatant.

The sample pretreatment for direct exosome detection from cell culture supernatant is not recommended for detection of exosomes from body fluids.

Specific sample pretreatment protocols are available for body fluids (plasma, urine) each optimized for its specific type of biological sample.

To ensure that detected exosomes originate from your cells of interest, culture the cells with exosome depleted fetal bovine serum (FBS), because normal FBS contains extremely high levels of exosomes that will contaminate the cell derived exosomes.



**Figure 3:** Cell Culture, human plasma and urine supernatant pretreatment workflow for direct exosome detection.

Please, refer to <http://www.immunostep.com/content/34-protocols> for additional technical information.

## PROTOCOL


### Isolate CD63<sup>+</sup> exosomes

1. Resuspend the capture beads by vortex for approximately 20 seconds.
2. Add 50µL of the capture beads to each 12x75mm Polystyrene Round Bottom tube (cytometer tube).
3. Add between 10-15µg of exosomes isolated by differential ultracentrifugation or until 100ul for direct exosomes. Previously prepared according to "Sample Preparation" to the appropriate tubes. Mix the reactions gently by pipetting up and down several times with a pipette and vortexing for few seconds.
4. Incubate in the dark overnight at room temperature (RT). NO STIRRING.
5. After overnight incubation wash the sample (bead-bound exosomes) by adding 1ml of Assay Buffer 1X.
6. Collect the magnetic beads by placing tubes on a magnetic rack and incubate 5 minutes or by centrifugation at 2500xg for 5 minutes.

- Remove supernatant from tubes by hand-decanting in the case of using the magnetic rack (Fig. 4) or by aspiration. Take care not to disturb the microspheres, and make sure not to leave more than 100ul of supernatant in the tube.

#### Stain exosomes for flow cytometry

- Add 5µL of the primary detection antibody to the bead-bound exosomes tube. Mix gently by pipetting and/or by tapping. It is advisable to prepare an additional tube with the appropriate isotype control or without exosomes, for background determination.
- Incubate in the dark 60 minutes at 2-8°C, without stirring.
- Wash the sample (bead-bound exosomes) by adding 1ml of Assay Buffer 1X.
- Collect the magnetic beads by placing tubes on a magnetic rack and incubate 5 minutes or by centrifugation at 2500xg for 5 minutes. Remove supernatant from tubes by hand-decanting in the case of using the magnetic rack (Fig. 4) or by aspiration. Take care not to disturb the microspheres, and make sure not to leave more than 100ul of supernatant in the tube.




**Figure 4:** Hand-decanting supernatant using a Magnetic Rack.

- Add 5 µL of the secondary detection antibody (5 µL) to the bead-bound exosomes tube. Mix the reactions gently by pipetting up and down several times with a micropipette.
- Incubate in the dark 30 minutes at 2-8°C, without stirring.
- Wash the sample (bead-bound exosomes) by adding 1 ml of Assay Buffer 1X.
- Collect the magnetic beads by placing tubes on a magnetic rack and incubate 5 minutes or by centrifugation at 2500xg for 5 minutes. Remove supernatant from tubes by hand-decanting decanting in the case of using the magnetic rack (Fig. 4) or by aspiration.
- Resuspend the sample in 350µL Assay Buffer 1X and acquire on a flow cytometer or store in the dark max up to 2 hours at 2-8°C, until the analysis is carried out.

#### ASSAY ACQUISITION

An adequate gating strategy FSC / SSC for 6 micron bead size and PerCP/APC, PerCP-Cy5/APC or PerCP-Cy5.5/APC helps bead population identification and discrimination of doublets on flow cytometer.


- Gate on the single population(s) on a Forward Scatter vs. Side Scatter plot in linear scale. (Fig. 5A).
- Gate on the single population(s) on a FL3 vs. FL4 channel (bead auto fluorescence) in logarithmic scale (Fig. 5B)
- Using the FL2 channels, determine whether or not any bead populations tested "positive" for the exosome. Note: A positive bead will produce a fluorescent peak in the FL2 channel.



**Figure 5:** Dot-plot gating strategy for acquisition and analysis. FSC vs SSC (A) and FL3 vs FL4 (B)

## FLOW CYTOMETRY ANALYSIS

ExoStep has been used for detection of exosome derived from different sources.



**Figure 6:** Flow analysis of exosomes bound to ExoStep. Cell culture exosomes, pre-enriched using Total Exosome Isolation from PC3 Cell Culture Media (A), human plasma (B) and urine (C) were resuspended in PBS and bound to CD63-capture beads during an overnight incubation. The following day the bead-bound exosomes were indirect stained with primary detection antibody (CD9-biotin/CD81-biotin) and streptavidin PE and analyzed by flow cytometry.

## REPRODUCIBILITY

Sample: PC3-derived exosomes.

Batch: 3 different batches.

Cytometer: Acquired every day in 2 different cytometers.

Replicate: 4 replicates assayed for 3 days not necessarily consecutive.

Protocol: Bead-Based Flow Cytometric Assays.

Intra assay: was determined calculating the deviation and the CV for each of the samples by batch. Was analyzed the mean of all typical deviations and CVs of 3 days for each lot. Finally was obtained the mean of the standard deviation and the CV of the three lots.


Inter assay: was determined the mean of the 4 repetitions for each day and compare them between each batch taking the standard deviation and the CV. Was calculated the mean deviation thus obtained and the CV of the three days.

	CV (%)
Intra Assay	10%
Inter Assay	11%

## PERFORMANCE DATA

Limit of Detection (LOD), dynamic range and linearity of exosome kit was assessed at Immunostep.

LOD is the lowest quantity of exosomes that is distinguished from the absence of analyte (a blank value), and as reference, was determined in  $>0,125\mu\text{g}$  which corresponds with  $1,25 \times 10^9$  vesicles. Whilst the upper limit or saturation level was established in  $10\mu\text{g}$ . For both technical specifications were used exosome form PC3 cell culture media ( $1 \times 10^6$  vesicles/ $\mu\text{l}$ ). (Fig. 7)



**Figure 7:** Dynamic range of the assay analyzed by flow cytometry. Relationship between background noise and specific signal at different exosome concentrations.

Several measurements of multiple concentrations of lyophilized exosomes were analyzed across the reportable range of the kit, finding the linearity of the kit in a broad range of concentrations, allowing fluorescence interpolation in the estimation of concentrations.

Please, refer to <http://www.immunostep.com/content/34-protocols> for additional technical information.

\* PE (Ex-Max 496 nm/Em-Max 578 nm); Excitation laser line 488 nm  
\* FITC (Ex-Max 494 nm/Em-Max 519 nm); Excitation laser line 488 nm  
\* CF-Blue (Ex-Max 401 nm/Em-Max 452 nm); Excitation laser line 405 nm

## REFERENCES

1. Yáñez-Mó M, Siljander P, Andreu Z, Bedina Zavec A, Borràs F, Buzas E et al. Biological properties of extracellular vesicles and their physiological functions. *Journal of Extracellular Vesicles*. 2015;4(1):27066.
2. Pitt JM, André F, Amigorena S, Soria JC, Eggermont A, Kroemer G, Zitvogel L. Dendritic cell-derived exosomes for cancer therapy. *J Clin Invest*. 2016.
3. Tkach M, Théry C. Communication by Extracellular Vesicles: Where We Are and Where We Need to Go. *2016 Cell* 10; 164(6):1226-32.
4. Becker A, Thakur BK, Weiss JM1, Kim HS, Peinado H, Lyden D Extracellular Vesicles in Cancer: Cell-to-Cell Mediators of Metastasis. *Cancer Cell* 2016 Dec 12;30 (6):836-848.
5. López-Cobo S, Campos-Silva C, Valés-Gómez M. Glycosyl-Phosphatidyl-Inositol (GPI)-Anchors and Metalloproteases: Their Roles in the Regulation of Exosome Composition and NKG2D-Mediated Immune Recognition. *Front Cell Dev Biol*. 2016 Sep 12;4:97.
6. Jonathan M. Pitt, Guido Kroemer, Laurence Zitvogel Extracellular vesicles: masters of intercellular communication and potential clinical interventions. *2016 J Clin Invest*. 2016;126(4):1139-1143
7. Campos S, Suárez H, Jara-Acevedo R, Linares-Espinós E, Martínez-Piñeiro L, Yáñez-Mó M, Valés-Gómez M. High sensitivity detection of extracellular vesicles immune-captured from urine by conventional flow cytometry. *Sci Rep*. 2019; Feb 14; 9 (1):2042.
8. Jara-Acevedo R, Campos-Silva C, Valés-Gómez M, Yáñez-Mó M, Suárez H, Fuentes M. Exosome beads array for multiplexed phenotyping in cancer. *J Proteomics*. 2019; Apr 30; 198: 87.
9. Théry C, Amigorena S, Raposo G, Clayton A. Isolation and Characterization of Exosomes from Cell Culture Supernatants and Biological Fluids. *Current Protocols in Cell Biology*. 2006.

## MANUFACTURED BY



**Immunostep S.L**  
Avda. Universidad de Coimbra, s/n  
Cancer Research Center (CIC)  
Campus Miguel de Unamuno  
37007 Salamanca (Spain)  
Tel. (+34) 923 294 827  
[www.immunostep.com](http://www.immunostep.com)