



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 



**Lifespan Technologies**  
PO Box 3352  
Salt Lake City, UT 84110  
Tel: 801-971-1818 Fax: 801-464-6116  
info@lifespantech.com

Product No: **K-1800**  
Low Molecular Weight Heparin  
ELISA for Buffer Samples  
Range: 0.03 – 10 µg/ml

## K-1800: Low Molecular Weight Heparin (LMWH) ELISA Kit

**INTENDED USE: THIS PRODUCT IS FOR RESEARCH USE ONLY. NOT INTENDED FOR CLINICAL OR DIAGNOSTIC USE.**

### **Kit includes:**

Heparin coated 96-well plate  
Detector -Enzyme Conjugate vial  
Conjugate Diluent  
TMB Solution  
Stop Solution  
Wash Concentrate 10X, (dilute 1 part plus 9 parts water)

### **Researcher must provide:**

Pipettes (8-channel multipipettor recommended)  
Absorbance microplate reader  
LMWH standards from USP reference or your heparin  
Tris Buffered Saline (TBS) pH 7.5 (10mM Tris 150mM NaCl)  
Plate Cover

### **Storage and Stability**

Kit can be stored unopened at 4°C for up to six months. Reconstituted detector enzyme conjugate is unstable and should be used immediately. If you wish to run less than a full plate, it should be stored as frozen aliquots at -80° C. Aliquots must be thawed **immediately** before use. After one thaw, any unused detector enzyme conjugate must be discarded. TMB solution should be protected from light.

### **Background**

Heparin is a glycosaminoglycan with alternating uronic acid and aminoglycoside units. It is an anticoagulant used either in its native unfractionated form (UFH) MW ~16 kD or in various partially depolymerized forms (LMWH) of 4-8 kD. The heparin-ELISA product number K-1800 is a quantitative enzyme-linked assay designed for the *in vitro* measurement of low molecular weight heparin levels in low protein content fluids such as buffer or urine. This assay measures heparin directly using a heparin binding protein which has been conjugated to HRP.

The heparin-ELISA is a competitive assay in which the colorimetric signal is inversely proportional to the amount of heparin present in the sample. Samples to be assayed are first mixed with the Detector-Enzyme Conjugate in wells of the heparin coated plate. Heparin in the sample competes with heparin bound to the plate for binding of the Detector-Enzyme Conjugate. The concentration of heparin in the sample is determined using a standard curve of known amounts of heparin.

### **Reagent Preparation**

**Heparin Standards:** Make dilutions of your heparin in Tris Buffered Saline (TBS) pH 7.5 (10mM Tris, 150mM NaCl) to obtain standards of 0.03, 0.1, 0.3, 1.0, 3.0 and 10.0 µg/mL. **Standardization should be performed using heparin that is the same heparin type contained in your unknowns.**

**Working Detector-Enzyme Conjugate:** Measure exactly **9.1 mL** of conjugate diluent and add to a clean tube. Perform a 'clean transfer' of the lyophilized Detector-Enzyme Conjugate into the 9.1 mL of conjugate diluent. This can be done by adding 500 microliters of the measured Diluent to reconstitute the Detector-Enzyme. Wait a minute to allow the lyophilized material to dissolve and then add the liquid back to the tube containing the conjugate diluent. Repeat this step two more times to be sure all the Detector-Enzyme Conjugate has been transferred from the vial to the tube. If you wish to perform less than a full plate, the reconstituted Detector-Enzyme Conjugate must be stored as aliquots at -80°C. Aliquots must be thawed **immediately** before use. After one thaw, any unused detector enzyme conjugate must be discarded.

**1X Wash Buffer:** Make a 1:10 dilution of 10X Wash Buffer in distilled or deionized water.

