

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten! See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere Liefer- und Versandbedingungen

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

linkedin.com/company/szaboscandic in





Instructions For Use

RA0325-C.5-IFU-RUO

Rev. Date: Dec. 15, 2014

Revision: 1

Page 1 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com

Tubulin-alpha (Microtubule Marker); Clone TU-16 (Concentrate)

Availability/Contents: <u>Item #</u> <u>Volume</u>
RA0325-C.5 <u>Volume</u>
0.5 ml

Description:

Species: Mouse

Immunogen: Full length native protein corresponding to Pig alpha-tubulin.

Clone: TU-16
Isotype: IgM, kappa
Entrez Gene ID: 7277 (Human)

Hu Chromosome Loc.: 2g35

Synonyms: Alpha-tubulin 1; Hum a tub1; Hum a tub2; Testis-specific alpha-tubulin; TUBA1A;

Tuba4a; Tubulin alpha-1 chain; Tubulin alpha-4A chain; Tubulin H2-alpha; Tubulin, alpha 1

(testis specific); Tubulin, alpha 4a; Tubulin, alpha, testis-specific

Mol. Weight of Antigen: 55kDa

Format: 200µg/ml of Ab purified from Bioreactor Concentrate. Prepared in 10mM PBS with 0.05% BSA

& 0.05% azide.

Specificity: This antibody recognizes an epitope of alpha-tubulin.

Background: Microtubules of the eukaryotic cytoskeleton perform essential and diverse functions and are

composed of a heterodimer of alpha- and beta-tubulin. The alpha-tubulin is a globular protein that exists in cells either as part of the soluble alpha/beta-tubulin dimer or polymerized into microtubules. In different species, it is coded by multiple tubulin genes that form tubulin classes (6 genes in human). Expressed tubulin genes are named tubulin isotypes. Some of the tubulin isotypes are expressed ubiquitously, while some have more restricted tissue expression. Alphatubulin is also subject to numerous post-translational modifications. Tubulin isotypes and their posttranslational modifications are responsible for multiple tubulin charge variants, or tubulin isoforms. Heterogeneity of alpha-tubulin is concentrated in the C-terminal structural domain.

Species Reactivity: Human, Pig, Cow, Mouse, Rat, Hamster, Chicken, and many others.

Positive Control: A431, HeLa, Raji, or NIH/3T3 cells. Colon tissue.

Cellular Localization: Cytoplasmic

Titer/ Working Dilution: Immunohistochemistry (Frozen and Formalin-fixed): 0.5-1 µg/ml

Flow Cytometry: 0.5-1 µg/million cells

 $\label{eq:mmunofluorescence:} Inmunofluorescence: 1-2 \ \mu g/ml \\ Western Blotting: 0.5-1 \ \mu g/ml$

Immunoprecipitation: 1-2 μg/500μg protein lysate

Microbiological State: This product is not sterile.

Storage: 2° C 8° C

ScyTek Laboratories, Inc. 205 South 600 West Logan, UT 84321 U.S.A.

CE

EmergoEurope (31)(0) 70 345-8570 Molsnstraat 15 2513 BH Hague, The Netherlands



Instructions For Use RA0325-C.5-IFU-RUO

Rev. Date: Dec. 15, 2014

Revision: 1

Page 2 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com

Uses/Limitations: Not to be taken internally.

For Research Use Only.

This product is intended for qualitative immunohistochemistry with normal and neoplastic formalin-fixed, paraffin-embedded

tissue sections, to be viewed by light

microscopy.

Do not use if reagent becomes cloudy. Do not use past expiration date.

Non-Sterile.

Ordering Information and Current Pricing at www.scytek.com

Procedure:

- Tissue Section Pretreatment (Required): Staining of formalin fixed, paraffin embedded tissue sections is significantly enhanced by pretreatment with Citrate Plus (ScyTek catalog# CPL500).
- Primary Antibody Incubation Time: We suggest an incubation period of 30 minutes at room temperature.
 However, depending upon the fixation conditions and the staining system employed, optimal incubation should be determined by the user.
- 3. **Visualization:** For maximum staining intensity we recommend the "UltraTek HRP Anti-Polyvalent Lab Pack" (ScyTek catalog# UHP125, see IFU for instructions) combined with the "DAB Chromogen/Substrate Bulk Pack (High Contrast)" (ScyTek catalog# ACV500, see IFU for instructions).

Precautions:

Contains Sodium Azide as a preservative (0.09% w/v).

Do not pipette by mouth.

Avoid contact of reagents and specimens with skin and mucous membranes.

Avoid microbial contamination of reagents or increased nonspecific staining may occur.

This product contains no hazardous material at a reportable concentration according to U.S. 29 CFR 1910.1200,

OSHA Hazardous Communication Standard and EC Directive 91/155/EC.

References:

1. Dráber, P., Draberova, E., Linhartova, I. and Viklicky, V. (1989) J. Cell Sci. 92, 519-528.

2. Dráber, P., Lagunowich, L.A., Draberova, E., Viklickjr, V. and Damjanov, I. (1988) Histochemistry 89, 485-492.

3. Dráber P, Leu FJ, Viklický V, Damjanov I. Immunohistochemical héterogeneity of alpha-tubulin in human epithelia revealed with monoclonal antibodies. Histochemistry. 1987;87(2):151-5.

Warranty:

No products or "Instructions For Use (IFU)" are to be construed as a recommendation for use in violation of any patents. We make no representations, warranties or assurances as to the accuracy or completeness of information provided on our IFU or website. Our warranty is limited to the actual price paid for the product. ScyTek Laboratories, Inc. is not liable for any property damage, personal injury, time or effort or economic loss caused by our products. Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining can cause inconsistent results. Variations in fixation and embedding or the inherent nature of the tissue specimen may cause variations in results. Endogenous peroxidase activity or pseudoperoxidase activity in erythrocytes and endogenous biotin may cause non-specific staining depending on detection system used.

Storage: 2° C

ScyTek Laboratories, Inc. 205 South 600 West Logan, UT 84321 U.S.A. CE

EC REP EmergoEurope (31)(0) 70 345-8570 Molsnstraat 15 2513 BH Hague, The Netherlands