



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

TMB Soluble Reagent (High Sensitivity)


Description: This liquid substrate for peroxidase consists of tetramethylbenzidine (TMB) plus dilute hydrogen peroxide in a single-reagent stabilized form. The reagent has been specifically formulated for measuring peroxidase in ELISA systems. This reagent is stable for long-term storage and provides sensitivity equal to, or greater than, that of OPD.

Form: 3,3',5,5'-tetramethylbenzidine

Contents: TMB in a dilute organic solvent with buffer, pH 6.0+ 0.5. Hydrogen peroxide is added at a concentration of 0.03%

Stability: Reagents are stable for at least 12 months when stored at room temperature, or 18 months when stored at 2-8°C. Avoid contamination of reagents with labware which has not been thoroughly cleaned. A slight yellow tinge may develop over time. This does not affect product performance. Do not use if solution darkens.

Uses/Limitations: Not to be taken internally.
For In-Vitro Diagnostic use.
Immunological applications.
Do not use if reagents become cloudy.
Do not use past expiration date.
Use caution when handling reagents.
Non-Sterile.



Ordering Information and Current Pricing at www.scytek.com


Availability:	Item #	Volume
	TM4125	125 ml
	TM4500	500 ml
	TM4999	1000 ml


Storage: Store at 2-8°C.

Precautions: Avoid contact with skin and eyes.
Harmful if swallowed.
Do Not pipette reagent by mouth.
Follow all Federal, State, and local regulations regarding disposal.

Activating Agents: Peroxidase

Light Sensitivity: Negligible for short exposure times

Storage: 2° C  8° C



ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.

CE 


EMERGO
Emergo Europe
Molenstraat 15
2513 BH Hague, The Netherlands


P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 – Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com

- Reaction Volume:** 50 - 100 ul per well in microtiter plates
- Reaction Time:** Approximately 15 minutes (Range 5 - 60 min.)
- Reaction pH:** Approximately pH 6.0 (Range 5.0 - 7.0)
- Reaction Temperature:** Room temperature
- Peak Wavelengths:** 650 nm, unstopped, blue reaction product
450 nm, stopped, yellow reaction product
- Stopping Solution:** Equal volume of Stop Buffer (cat# TSB). Stopped reactions show increased absorbance values of approximately 2-fold over unstopped reactions.
- Reaction Stability:** Stopped reactions are stable for at least 30 minutes to several hours depending on the level of peroxidase activity. Intense reactions may precipitate on prolonged standing. This can be prevented by increasing concentration of stopping solution.

References:

1. Kusano-Arai O, Fukuda R, Kamiya W, Iwanari H, Hamakubo T. Kinetic exclusion assay of monoclonal antibody affinity to the membrane protein Roundabout 1 displayed on baculovirus. *Analytical biochemistry*. 2016 Jul 1;504:41-9.
2. Pont-Thibodeau D, Robitaille N, Gauvin F, Thibault L, Rivard GÉ, Lacroix J, Tucci M. Incidence of hypotension and acute hypotensive transfusion reactions following platelet concentrate transfusions. *Vox sanguinis*. 2016 Feb 1;110(2):150-8.
3. Ogino, K., Murakami, I., Wang, DH., Tsukiyama, Y., Takahashi, H., Kubo, M., Sakano, N., Setiawan, H., Bando, M., Ohmoto, Y., (2013) Evaluation of serum arginase I as an oxidative stress biomarker in a healthy Japanese population using a newly established ELISA. *Clinical Biochemistry*, Volume 46, Issues 16-17, November 2013, pages 1717-1722.
4. Dumont, N., Aubin, E., Proulx, D. P., Lemieux, R. and Bazin, R. (2009), Increased secretion of hyperimmune antibodies following lipopolysaccharide stimulation of CD40-activated human B cells *in vitro*. *Immunology*, 126: 588–595. doi: 10.1111/j.1365-2567.2008.02915.x
5. Kato, S., Itoh, K., Ochiai, M., Iwai, A., Park, Y., Hata, S., Takeuchi, K., Ito, M., Imaki, J., Miura, S., Yakabi, K. and Kobayashi, M. (2008), Increased pentosidine, an advanced glycation end-product, in urine and tissue reflects disease activity in inflammatory bowel diseases. *Journal of Gastroenterology and Hepatology*, 23: S140–S145. doi: 10.1111/j.1440-1746.2008.05552.x
6. Yokota, H., Hiramoto, M., Okada, H., Kanno, Y., Yuri, M., Morita, S., Naitou, M., Ichikawa, A., Kato, M., Suzuki, H., (2007) Absence of Increased α 1-Microglobulin in IgA Nephropathy Proteinuria. *Molecular & Cellular Proteomics*, April 2007, 6: pages 738-744. doi: 10.1074/mcp.M600336-MCP200
7. Yano, K., Tsuda, E., Washida, N., Kobayashi, F., Goto, M., Harada, A., Ikeda, K., Higashio, K. and Yamada, Y. (1999), Immunological Characterization of Circulating Osteoprotegerin/Osteoclastogenesis Inhibitory Factor: Increased Serum Concentrations in Postmenopausal Women with Osteoporosis. *J Bone Miner Res*, 14: 518–527. doi: 10.1359/jbmr.1999.14.4.518

Storage: 2° C  8° C

 ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.



Emergo Europe
Molenstraat 15
2513 BH Hague, The Netherlands