

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten! See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere Liefer- und Versandbedingungen

## Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

linkedin.com/company/szaboscandic in





# Instructions For Use UAM-IFU

Rev. Date: June 13, 2008

Revision: 1

Page 1 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Fax (435) 755-0015 - www.scytek.com

## UltraTek Alk-Phos Anti-Mouse Lab Pack

Species of Origin: Goat

Antigen Specificity: Anti-Mouse (Mouse and Rat ).

Preadsorbed Against: Human

**Enzyme Conjugate:** Alkaline Phosphatase **Chromogen Substrate:** None Provided

**Uses/Limitations:** For In-vitro use only.

Do not use past expiration date. For immunohistochemical studies.

#### Availability:

REF # Volume

UAM125 125ml Super Block, 125ml UltraTek Anti-Mouse, 125ml UltraTek Alk-Phos.

UAM500 500ml Super Block, 500ml UltraTek Anti-Mouse, 500ml UltraTek Alk-Phos.

UAM999 1000ml Super Block, 1000ml UltraTek Anti-Mouse, 1000ml UltraTek Alk-Phos.

**Storage:** 2-8º Centigrade.

#### Procedure:

- 1. Deparaffinize and rehydrate tissue section.
- Wash 2 times in Tris Buffered Saline + Tween 20 (20X) pH 7.4 (catalog # TBT500).
- If required, incubate tissue in digestive enzyme (catalog # PSS060 or TSS155) or Citrate Plus (catalog # CPL500).
- 4. Wash 3 times in Tris Buffered Saline + Tween 20 (20X) pH 7.4 (catalog # TBT500).
- Apply Super Block and incubate for 5 minutes at room temperature to block nonspecific background staining.
   Note: Do not exceed 10 minutes or there may be a reduction in desired stain.
- Wash 1 time in Tris Buffered Saline + Tween 20 (20X) pH 7.4 (catalog # TBT500).
- 7. Apply primary antibody and incubate according to manufacturer's protocol.
- 8. Wash 3 times in Tris Buffered Saline + Tween 20 (20X) pH 7.4 (catalog # TBT500).
- Apply UltraTek Anti-Mouse (yellow solution), and incubate for 10 minutes at room temperature.
- 10. Wash 3 times in Tris Buffered Saline + Tween 20 (20X) pH 7.4 (catalog # TBT500).
- 11. Apply UltraTek Alk-Phos (red solution), and incubate for 10 minutes at room temperature.
- 12. Wash 3 times in Tris Buffered Saline + Tween 20 (20X) pH 7.4 (catalog # TBT500).
- 13. Apply chromogen intended for use with Alkaline Phosphatase and incubate as desired.
- 14. For optimal results counterstain using Hematoxylin for Automation (catalog # HAQ500).
- 15. Coverslip using mounting media of choice (catalog # AMT030 or PMT030).

The material contained herein is the property of ScyTek Laboratories, Inc. and is confidential. Unauthorized use or copying is prohibited.



# Instructions For Use UAM-IFU

Rev. Date: June 13, 2008

**Revision: 1** 

Page 2 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Fax (435) 755-0015 - www.scytek.com

#### **Troubleshooting Guide**

#### Overstaining:

- 1. Concentration of the primary antibody was too high or the incubation time was too long.
- 2. Temperature during incubation was too high.
- 3. Incubation time with link antibody or streptavidin/enzyme label was too long.

#### **Nonspecific Background Staining:**

- 1. Rinsing between steps was inadequate.
- 2. Tissue was allowed to dry with reagents on.
- 3. Folds in tissue trapped reagents.
- 4. Tissue contains endogenous alkaline phosphatase.
- 5. Tissue contains endogenous biotin.
- 6. Antigen migrated in tissue.
- 7. Excessive tissue adhesive on slides.
- 8. Inadequate blocking with protein block.

#### Weak Staining:

- 1. Primary antibody concentration was too low or incubation time was too short.
- 2. Reagents are past their expiration date.
- 3. Inadequate removal of wash water between steps, resulting in dilution of reagents.
- 4. Counterstain or mounting media were incompatible and dissolved the chromogen reaction product.
- 5. Room temperature was excessively cool.
- 6. The primary antibody does not recognize an antigen that survives fixation and embedding in high enough amounts.
- 7. Excessive incubation with protein block (Super Block).

#### No Staining:

- 1. Steps were inadvertently left out.
- 2. There is no antigen in the tissue.
- 3. The primary antibody is not of mouse or rat origin.
- 4. Chromogenic substrate has been replaced with another that is not intended for use with alkaline phosphatase.

**Precautions:** Handle with care and dispose of according to all regulations.