



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC Handels GmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

UltraTek HRP Anti-Polyvalent Lab Pack

Description: The UltraTek staining kit provides unmatched sensitivity with incubation times of 10 minutes each for the Link Antibody and Enzyme Label. The bulk kits are ideal for high volume laboratories. Each Pack contains one bottle of Super Block (universal protein block), one bottle of Biotinylated Antibody (Polyvalent), and one bottle of Horseradish Peroxidase Labeled Streptavidin. These Lab-Packs provide an extremely economical alternative for automated staining systems and we encourage you to evaluate the addition of this in your current system.


Species of Origin: Goat
Antigen Specificity: Anti-Polyvalent (Mouse, Rat, Rabbit and Guinea Pig).
Preadsorbed Against: Human
Enzyme Conjugate: Horseradish Peroxidase
Chromogen Substrate: None Provided

Contains: One container of Super Block.
 One container of Anti-polyvalent.
 One container of HRP.

Uses/Limitations: Not to be taken internally.
 For In-Vitro Diagnostic use.
 Histological applications.
 Do not use if reagents become cloudy.
 Do not use past expiration date.
 Use caution when handling reagents.
 Non-Sterile.

Control Tissue: Any FFPE tissue.
 Any Fresh or Frozen tissue.
 Cell smear or spin.

Ordering Information and Current Pricing at www.scytek.com




Availability:	<u>Item #</u>	<u>Volume</u>
	UHP125	125ml each
	UHP500	500ml each
	UHP999	1000ml each


Storage: Store at 2-8°C.

Precautions: Avoid contact with skin and eyes.
 Harmful if swallowed.
 Follow all Federal, State, and local regulations regarding disposal.

Storage: 2° C  8° C

 ScyTek Laboratories, Inc.
 205 South 600 West
 Logan, UT 84321
 U.S.A.

CE 


 Emergo Europe
 Molenstraat 15
 2513 BH Hague, The Netherlands

Recommended, But Not Included:


Item #	Description
PBE500	Phosphate Buffered Saline + Tween 20 (10x) pH 7.4
or TBE500	Tris Buffered Saline + Tween 20 (10x) pH 7.5
CPL500	Citrate Plus
ADA500	Peroxide Block for Image
ACT500	DAB Chromogen/Substrate Kit (High Contrast)
HMM500	Hematoxylin, Mayer's (Lillie's Modification)
BRT500	Bluing Reagent


Procedure:

- Rehydrate tissue slides.
- In a glass or plastic (Autoclavable) Coplin jar, add 5 ml of Citrate Plus (CPL) and 45 ml of deionized water. *(Not included)*
- Submerge slides in diluted Citrate Plus and loosely cap.
- Add Distilled water to bottom of Autoclave or Pressure Cooker (about 1 inch deep in Pressure Cooker).
- Place Coplin jar in Pressure Cooker or Autoclave.
- Turn heat on and allow pressure to rise to 20-25 PSI.
- Maintain pressure at 20-25 PSI for 5 minutes.
- Turn off heat source and allow to cool.
- When pressure has dropped to ambient, carefully remove lid or open door.
- Using tongs, remove Coplin Jar and place on counter.
- Once Coplin Jar cools to room temperature remove slides, rinse several times in buffer and proceed with staining as usual.
- Apply Peroxide Block for Image Analysis (ADA) and incubate slide for 10-15 minutes. *(Not included)*
- Rinse 3 times in buffer.
- Apply Super Block (AAA), and incubate for 5 minutes at room temperature to block nonspecific background staining. **Note:** Do not exceed 10 minutes or there may be a reduction in desired stain.
- Rinse 3 times in buffer.
- Apply primary antibody and incubate according to manufacturer's protocol.
- Rinse 3 times in buffer.
- Apply UltraTek Anti-Polyvalent (ABN) and incubate for 10 minutes at room temperature.
- Rinse 3 times in buffer.
- Apply UltraTek HRP (ABL) and incubate for 10 minutes at room temperature.
- Rinse 3 times in buffer followed by 1 rinse in DI water.


WARNING: DAB is a suspected carcinogen. Handle with care and dispose of according to all regulations.

- Add 1 drop (40-50ul) DAB Chromogen (ACB) to each 1ml of DAB Substrate High Contrast (ACU), mix by swirling and apply to tissue for 5 minutes. *(Not included)*
- Rinse 1 time in DI water.

Storage: 2° C  8° C

 ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.


Emergo Europe
Molenstraat 15
2513 BH Hague, The Netherlands

24. Apply DAB Chromogen/Substrate mixture and incubate for a second 5 minute period.
25. Rinse 3 times in buffer.
26. Apply Hematoxylin, Mayer's (HMM) and incubate for 1 minute. *(Not included)*
27. Rinse 3 times in distilled water.
28. Apply Bluing Reagent (BRT) and incubate for 5-10 seconds. *(Not included)*
29. Rinse immediately in distilled or deionized water.
30. Dehydrate slides and clear in xylene or xylene substitute.
31. Coverslip using a permanent mounting media. *(Not included)*

Troubleshooting Guide

Overstaining:


1. Concentration of the primary antibody was too high or the incubation time was too long.
2. Temperature during incubation was too high.
3. Incubation time with link antibody or streptavidin/enzyme label was too long.


Nonspecific Background Staining:

1. Rinsing between steps was inadequate.
2. Tissue was allowed to dry with reagents on.
3. Folds in tissue trapped reagents.
4. Tissue contains endogenous peroxidase.
5. Tissue contains endogenous biotin.
6. Antigen migrated in tissue.
7. Excessive tissue adhesive on slides.
8. Inadequate blocking with protein block.


Weak Staining:

1. Primary antibody concentration was too low or incubation time was too short.
2. Reagents are past their expiration date.
3. Inadequate removal of wash water between steps, resulting in dilution of reagents.
4. Counterstain or mounting media were incompatible and dissolved the chromogen reaction product.
5. Room temperature was excessively cool.
6. The primary antibody does not recognize an antigen that survives fixation and embedding in high enough amounts.
7. Excessive incubation with protein block (Super Block).

Storage: 2° C  8° C

 ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.


Emergo Europe
Molenstraat 15
2513 BH Hague, The Netherlands

References:

1. Vahidinia Z, Alipour N, Atlasi MA, Naderian H, Beyer C, Azami Tameh A. Gonadal steroids block the calpain-1-dependent intrinsic pathway of apoptosis in an experimental rat stroke model. *Neurological Research*. 2016 Nov 11:1-1.
2. Zilber AL, Belli P, Artois M, Kodjo A, Djelouadi Z. First Observation of *Leptospira interrogans* in the Lungs of *Rattus norvegicus*. *BioMed Research International*. 2016 Oct 5;2016.
3. Lee YM, Kadiwal S, Lee KH, Park JC, Jeong WJ, Choi KY. Sur8 mediates tumorigenesis and metastasis in colorectal cancer. *Experimental & Molecular Medicine*. 2016 Jul 1;48(7):e249.
4. Bonnefont-Rebeix C, Fournel-Fleury C, Ponce F, Belluco S, Watrelot D, Bouteille SE, Rapiteau S, Razanajaona-Doll D, Pin JJ, Leroux C, Marchal T. Characterization of a novel canine T-cell line established from a spontaneously occurring aggressive T-cell lymphoma with large granular cell morphology. *Immunobiology*. 2016 Jan 31;221(1):12-22.
5. Chimenti C, Scopelliti F, Vulpis E, Tafani M, Villanova L, Verardo R, De Paulis R, Russo MA, Frustaci A. Increased oxidative stress contributes to cardiomyocyte dysfunction and death in patients with Fabry disease cardiomyopathy. *Human pathology*. 2015 Nov 30;46(11):1760-8.
6. Balemans MC, Ansar M, Oudakker AR, van Caam AP, Bakker B, Vitters EL, van der Kraan PM, de Bruijn DR, Janssen SM, Kuipers AJ, Huibers MM. Reduced Euchromatin histone methyltransferase 1 causes developmental delay, hypotonia, and cranial abnormalities associated with increased bone gene expression in Kleeftstra syndrome mice. *Developmental biology*. 2014 Feb 15;386(2):395-407.
7. Hascalik S, Celik O, Tagluk ME, Yildirim A, Aydin NE. Effects of highly purified urinary FSH and human menopausal FSH on uterine myoelectrical dynamics. *Molecular human reproduction*. 2010 Mar 1;16(3):200-6.
8. Castillo SS. A possible role of insulin-like growth factor-II C-peptide in regulating the function of steroidogenic cells in adult frog adrenal glands. *Acta histochemica*. 2008 Nov 30;110(6):451-61.
9. Saglam E, Ates LE, Kayir H, Celik T, Terzioğlu B, Uzbay T. Escitalopram increases cortical nitric oxide synthase (NOS) in rat brain during ethanol withdrawal. *Nitric Oxide*. 2008 Nov 30;19(3):284-8.
10. Celik O, Hascalik S, Elter K, Tagluk ME, Gurates B, Aydin NE. Combating endometriosis by blocking proteasome and nuclear factor-κB pathways. *Human reproduction*. 2008 Nov 1;23(11):2458-65.
11. Feng J, El-Assal ON, Besner GE. Heparin-binding epidermal growth factor-like growth factor reduces intestinal apoptosis in neonatal rats with necrotizing enterocolitis. *Journal of pediatric surgery*. 2006 Apr 30;41(4):742-7.
12. Blandamura S, Marioni G, de Filippis C, Giacomelli L, Segato P, Staffieri A. Temporal bone and sinonasal inverted papilloma: the same pathological entity?. *Archives of Otolaryngology–Head & Neck Surgery*. 2003 May 1;129(5):553-6.

Storage: 2° C



8° C


 ScyTek Laboratories, Inc.
 205 South 600 West
 Logan, UT 84321
 U.S.A.


IVD

EC REP

 Emergo Europe
 Molenstraat 15
 2513 BH Hague, The Netherlands