



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

## Product Description

**Catalog Number:** 354488

**Product Description:** Corning® Biocoat™ 175cm<sup>2</sup> straight neck flask, clear, gelatin coated, with filter vent cap

**Component Materials:**

Gelatin	-	Porcine Gelatin
Flask	-	Polystyrene
Cap	-	Polyethylene with color concentrate
Filter	-	Versapore acrylic copolymer membrane

**Bioburden:**

This lot has been tested and found negative for Bacteria and Fungi.

**Cell Attachment and Growth Characteristics:**

This lot has been tested for the attribute of Cell Attachment and Growth utilizing HUVEC cells, an attachment-dependent mammalian cell line in a serum free media.

**Gelatin:**

The flasks are coated with Porcine Gelatin via a proprietary method, which creates a uniform, dense gelatin network on the plastic surface.

**Storage Conditions:**

For optimal performance, the unopened product should be stored in a dry place at 4-30 °C. (Do not freeze)

**Handling:**

The product is ready to use and does not require washing of the coated surface prior to use.

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual and microscopic examination of the product.

Bacteria & Fungi Free : Turbidity test

Packaging: Inspection for seal and barrier integrity, accurate labeling and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian Date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 1