



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 3291

Product Description: Corning® CellBIND® surface, 150 cm² cell culture flask, canted neck, anti-tip skirt, vent cap

Component Materials:

- Flask - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
- Cap - Virgin High Density Polyethylene, meets *USP, Class VI* requirements for plastic containers and closures. Heavy metal free (meets *CONEG* req.) color concentrate. 0.2 µM microporous PTFE membrane, meets *USP, Class VI* requirements for plastic containers and closures.

Product Dimensions:

- | | | | | | |
|--------------|---|----------|----------------|---|--------------|
| Length w/cap | - | 8.93 in. | Height (depth) | - | 1.54 in. |
| Width | - | 4.40 in. | Tolerances | - | +/- 0.05 in. |

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.

Sterility Assurance Level: SAL 10⁻⁵

Pyrogens:

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72: *Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing*. Results: less than 0.1EU/mL.

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

Surface Characterization:

Surface is characterized to be hydrophilic and negatively charged, composed of ≥20% oxygen atoms. This surface composition has been developed for cell attachment and growth.

Cell Attachment and Growth Characteristics:

The product has been tested for the attribute of cell attachment and growth utilizing an attachment- dependent mammalian cell line in a serum supplemented media.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

- | | |
|-------------------------|---|
| Integrity Testing: | Forward pressurization of the product to 10 inches of water column. |
| Cell Culture Treatment: | Wettability test using deionized distilled water to insure the presence of a hydrophilic surface. |
| Visual Attributes: | Visual and microscopic examination of the product. |
| Packaging: | Inspection for seal and barrier integrity, accurate labeling, and correct product configuration. |

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Revision Date:

03-01-14

Rev No: 9