



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 3384

Product Description: Corning® HTS Transwell® 96 Well, TC Treated, 8.0 µm pore size, with barcode label, in a receiver plate with lid

Component Materials:

- Receiver Plate - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
- Lid - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
- Transwell body - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
- Filter - Transparent polyester membrane, meets *USP, Class VI* requirements for plastic containers and closures.
- Adhesive - Proprietary Acrylate, meets *ISO 10993* Biocompatibility requirements.
- Barcode Label: - White polypropylene base, acrylic adhesive

Special Format:

Barcode label (format 128) on A1 short side

Product Dimensions:

- | | | | | | |
|-----------------------------|---|------------------------|--|---|-----------------------|
| Length of Plate | - | 5.030 in. | Bottom of Transwell to receiver plate | - | .050 in. |
| Width of Plate | - | 3.365 in. | Apical volume added/ Transwell per well | - | 75µL |
| Height in a reservoir w/lid | - | .770 in. | Bilateral volume added per receiver plate well | - | 235µL |
| Tolerances of Dimensions | - | +/- .010 in. | Suggested working volume | - | 75µL (A) to 235µL (B) |
| Maximum volume | - | 100µL (A) to 290µL (B) | | | |

Sterilization:

The lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*

Sterility Assurance Level: SAL 10⁻³

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

- Visual Attributes: Visual and microscopic examination of the product.
- Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.
- Cell Culture Treatment of Membrane: Wettability test using water to insure the presence of a hydrophilic surface.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 2