



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

## Product Description

**Catalog Number:** 3545

**Product Description:** Corning® Disposable 1L Spinner Flask, Side Arm Accessory, with MPC quick connect, GL45 thread, 0.2um membrane

**Component Materials:**

- Cap - Virgin High Density Polyethylene, meets *USP, Class VI* requirements for plastic containers and closures. Heavy metal free (meets CONEG requirements) color concentrate.
- Filter - 0.2µm PTFE membrane, meets *USP, Class VI* requirements for plastic containers and closures.
- Tubing - C-Flex thermoplastic elastomer, *meets USP, Class VI* requirements for plastic containers and closures
- MPC Quick Connect - Medical Grade PC, *meets USP, Class VI* requirements.

**Product Dimensions:**

- |              |   |              |            |   |               |
|--------------|---|--------------|------------|---|---------------|
| Cap Diameter | - | 45mm(1.77in) | Cap Height | - | 25.3mm(1.0in) |
| Tolerances   | - | +/- .050 in. |            |   |               |

**Sterilization:**

The lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (AAMI TIR 27) *Sterilization of health care products-Requirements for validation and routine control-Radiation sterilization*  
Sterility Assurance Level: SAL 10<sup>-6</sup>

**Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:**

This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacture; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 3