



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Corning Incorporated Life Sciences

Registered
ISO 9001:2008

Product Description

Catalog Number: 3843

Product Description: Corning® 96-well, Poly-D-Lysine Coated, white plate with clear flat bottom, with lid

Component Materials:

- Plate walls - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures. White concentrate
- Plate bottom - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
- Lid - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
- Poly-D-Lysine - Synthetic or manufactured materials

Product Dimensions:

Length of Plate	-	5.030 in.	Diameter of Well @ top	-	.270 in.
Width of Plate	-	3.365 in.	Diameter of Well @ bottom	-	.250 in.
Depth of Well	-	.420 in.	Height with Lid	-	.650 in.
Tolerances of Dimensions	-	+/- .010 in.	Volume per Well	-	.36 mL

Bioburden:

This product has been tested for Bioburden contamination. Results: Less than .52% of the wells contaminated.

Pyrogens:

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72: 2002/ (R) 2010: *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternatives to batch testing*. Results: ≤ 0.5 EU/mL (≤ 20 EU/device)

Poly-D-Lysine:

The plates are coated with a 70-150 KD PDL polymer via a proprietary method, which creates a uniform net positive charge on the plastic surface. Manufactured wholly from synthetic or manufactured materials and does not contain any raw materials produced from or substances derived of animal origin.

Cell Attachment and Growth Characteristics:

The product has been tested for the attribute of cell attachment and growth utilizing an attachment-dependent mammalian cell line in a serum free media.

Optical Characteristics:

The product is made of opaque white polystyrene walls to minimize well to well crosstalk and background fluorescence and /or luminescence. The bottom is made of clear polystyrene to permit direct microscopic viewing.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual and microscopic examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling and correct product configuration.

Cell Growth:

Visual for cell attachment.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian Date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 4