



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 430490

Product Description: Corning® 4 mL cryogenic vial, internally threaded, round bottom

Component Materials:

- Vial - Virgin Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.
- Cap - Virgin Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures. Heavy metal free (meets *CONEG* req.) color concentrate.
- Washer - Silicone, meets *USP, Class VI* requirements for plastic containers and closures.

Product Dimensions:

- | | | | | | |
|----------------------|---|----------|-----------------|---|--------------|
| Length of vial w/cap | - | 2.71 in. | Diameter of cap | - | 0.47 in. |
| Diameter of vial | - | 0.50 in. | Tolerances | - | +/- 0.05 in. |

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (TIR 33) *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*

Sterility Assurance Level: SAL 10⁻⁵

RNase/DNase Testing:

This product has been tested and is free of any detectable RNase/DNase contamination.

Pyrogens:

The product has been tested and has met the criteria established in ANSI/AAMI ST 72:2002/(R) 2010: *Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing.* Results: less than 0.1 EU/mL

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Run number for that year.

Rev No: 7