



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 9018

Product Description: Costar® 96-well, high binding assay plate, flat bottom, without lid

Component Materials:
Plate - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.

Product Dimensions:

Length of Plate	-	5.030 in.	Diameter of Well @ top	-	.270 in.
Width of Plate	-	3.365 in.	Diameter of Well @ bottom	-	.250 in.
Depth of Well	-	.420 in.	Height	-	.560 in.
Tolerances of Dimensions	-	+/- .010 in.	Working Volume per well	-	.36 mL

Surface Characterization:

Surface is characterized to be high binding. Binding capacity is approximately 500 ng IgG/cm².

Optical Characteristics:

The product is made of clear polystyrene to allow background absorbance to be \pm .005 optical density units from the mean.

Binding Characteristics:

Each lot has been tested for the attribute of well to well protein binding utilizing a modified ELISA to measure precision and percent low and high wells from the mean. Coefficient of variation must be \leq 3% and high and low wells \leq 8% from the mean. Product has a Costar Certified Surface Chemistry.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:	Visual and microscopic examination of the product.
Packaging:	Inspection for accurate labeling, and correct product configuration.
Functional:	Tested for protein binding precision.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 7